

INFORMES AQUAS

January 2017

Catalan Arthroplasty Register. Third report (2005-2014)



Generalitat de Catalunya
Departament de Salut



Agència de Qualitat i
Avaluació Sanitàries de Catalunya

The Agency for Health Quality and Assessment of Catalonia (AQuAS) is a public entity of the Catalan Health Ministry of the Government of Catalonia. Its mission is generating relevant knowledge to contribute to the improvement of the quality, safety and sustainability of the Catalan Health Care System and thus easing the decision-making process for citizens and health care managers and professionals. The Agency is a founding member of the International Network of Agencies of Health Technology Assessment (INAHTA) and of the International School on Research Impact Assessment (ISRIA), it is a corporative member of the Health Technology Assessment International (HTAi), of the CIBER group (Networked Biomedical Research Centers) in Epidemiology and Public Health, of the Rapid Assessment and Research (RAR) group in Health Care Services and Results, acknowledged by the Catalan Government, and the REDISSEC (Spanish Research Network on Health Care in Chronic Diseases).

It is recommended that this document be cited as follows: Domingo L, Arias J, Martínez O, Espallargues M. On behalf of the Executive Committee and the Advisory Committee of the RACat. Catalan Arthroplasty Register. Third report (2005-2014). Barcelona: Agency for Health Quality and Assessment of Catalonia. Catalan Health Ministry. Government of Catalonia; 2017.

Any parties interested in this document should contact:

Agency for Health Quality and Assessment of Catalonia. Roc Boronat, 81-95 (second floor).

08005 Barcelona (Spain). Tel.: +34 93 551 3888 | Fax: +34 93 551 7510 | <http://aquas.gencat.cat>

© 2017, Generalitat de Catalunya. Agency for Health Quality and Assessment of Catalonia

Published by: Agency for Health Quality and Assessment of Catalonia.

1st edition: Barcelona, January 2017

Translation: Diarmuid Magourty

Designed by: Atona



The Agency has full copyright of this document, which can be publicly reproduced, distributed and communicated, in full or in part, by any means, as long as it is not used for commercial ends and its authors and origin are explicitly cited. The complete licence can be consulted at: <http://creativecommons.org/licenses/by-nc-nd/3.0/legalcode>

Catalan Arthroplasty Register. Third report (2005-2014)

Authorship

Laia Domingo Torrell^{1,2}, Jorge Arias de la Torre^{1,3}, Olga Martínez Cruz¹, Mireia Espallargues Carreras^{1,2}

On behalf of the Executive Committee and the Advisory Committee of the RACat.

Affiliation:

1. Agència de Qualitat i Avaluació Sanitàries de Catalunya (AQuAS)
2. Red de Investigación en Servicios de Salud en Enfermedades Crónicas (REDISSEC)
3. Consorcio de Investigación Biomédica en Red de Epidemiología y Salud Pública (CIBERESP)

Declaration of conflict of interests: The authors hereby state that they do not have any conflicts of interest in relation to this document.

ACKNOWLEDGEMENTS

The research team of the Agency for Health Quality and Assessment of Catalonia (AQuAS, as per the acronym in Catalan), which has coordinated the project and written this annual report, would like to thank all the disinterested collaboration of professionals from 53 participating centres of RACat, for their efforts in ensuring the continuity of the project. Also, the research team would also like to thank the manufacturing and distribution companies that provided valuable information on their products; and finally, to specially thank the members of the Advisory Committee of the RACat for reviewing the document. The table below lists the members of the Executive Committee and the Advisory Committee of the RACat.

| Members of the Executive Committee and the Advisory Committee of the RACat 2011 | |
|--|---------------------|
| Position and centre | Member |
| Executive Committee | |
| Subdirector - CatSalut | Josep Maria Argimón |
| Chairman of the SCOTT, Head of the 'Septic Pathology of the Musculoskeletal System and Trauma' Section at the OST Service of Parc de Salut Mar | Lluís Puig |
| Director of the AQuAS | Antoni Dedeu |
| Advisory Committee | |
| Chairman of the Advisory Committee, head of OST of the Vall d'Hebron Hospital Master Plan of Rheumatic Diseases and the Locomotive System | Joan Nardi |
| SCOTT Members | |
| Head of OST, Hospital de Mataró | Jaume Auleda |
| Clinical Director of COT, Hospital Vall d'Hebron | Enric Cáceres |
| Head of OST, H.U. Bellvitge | Federico Portabella |
| Head of OST, H.U. Vall d'Hebron | Joan Nardi |
| Head of OST, H.U. Dr. Josep Trueta de Girona | Miguel A Froufe |
| Head of OST, H.U. Santa Maria de Lleida | Francesc Pallisó |
| ICEMEQ Director, Hospital Clínic de Barcelona | Santiago Suso |
| CatSalut Members | |
| Management of Health Care Services Purchase and Assessment | Cristina Nadal |
| Waiting lists coordinator | Ignasi Ruano |
| Head of Activity Registries Management Division | Montse Bustins |
| Head of Health Care Services Purchase Division | Carme Casas |
| Management of the Barcelona Health Care Consortium | Jaume Estany |
| AQuAS | |
| Head of Quality and Assessment Department, AQuAS | Mireia Espallargues |
| Researcher, AQuAS | Laia Domingo |
| Administrative assistant-technician, AQuAS | Olga Martínez Cruz |

TABLE OF CONTENTS

| | |
|--|----|
| 1. Introduction | 15 |
| 2. Objectives | 16 |
| 3. Methodology | 17 |
| 3.1. Study population | 17 |
| 3.2. Data collection and sources of information | 17 |
| 3.3. Data protection and confidentiality | 18 |
| 3.4. Quality control and data validation | 19 |
| 3.5. Primary arthroplasty classification system | 19 |
| Primary knee arthroplasties | 19 |
| Primary hip arthroplasties | 20 |
| 3.6. Statistical analysis | 20 |
| Completeness and quality of the information | 20 |
| Descriptive analyses of patients and arthroplasties | 21 |
| Implant survival analysis | 22 |
| 4. Knee Arthroplasties results | 23 |
| 4.1. Scope and quality of data | 23 |
| 4.2. Demographic and patient treatment process profile | 30 |
| 4.3. Characteristics of primary arthroplasties | 35 |
| 4.4. Characteristics of implant models | 40 |
| 4.5. Implant survival in knee arthroplasties | 45 |
| Follow-up description | 45 |
| Cumulative revision rate | 46 |
| 5. Hip Arthroplasties results | 52 |
| 5.1. Scope and quality of the data | 52 |
| 5.2. Demographic and patient treatment process profile | 58 |
| 5.3. Characteristics of primary arthroplasties | 63 |
| 5.4. Characteristics of implant models | 68 |
| 5.5. Implant survival in hip arthroplasties | 72 |

| | |
|--|----|
| Follow-up description..... | 72 |
| Cumulative revision rate | 73 |
| 6. Discussion | 79 |
| 6.1. Quality of the data: volume of interventions and completeness | 79 |
| 6.2. Description of patients and the healthcare process..... | 79 |
| 6.3. Characteristics of primary arthroplasties and implant models | 80 |
| 6.4. Implant survival..... | 80 |
| 6.5. Strengths and weaknesses..... | 81 |
| 7. Final Remarks and Next Steps..... | 82 |
| 7.1. Latest achievements..... | 82 |
| 7.2. Next steps..... | 82 |
| 8. References | 84 |

LIST OF TABLES

| | |
|--|----|
| Table 1. Variables collected from RACat's source information | 18 |
| Table 2. Classification system for primary hip and knee arthroplasties | 20 |
| Table 3. Number of knee arthroplasties by health region period and by type of arthroplasty .. | 25 |
| Table 4. Number of knee arthroplasties by health care region, centre, period and by type of arthroplasty | 25 |
| Table 5. Percentages of classified primary knee arthroplasties in relation to the volume sent to the RACat..... | 29 |
| Table 6. Number of knee arthroplasties with the operated side (left or right) reported by period | 29 |
| Table 7. Age and sex of patients undergoing a primary total knee arthroplasty by health care regions and period | 31 |
| Table 8. Age and sex of patients undergoing a primary partial knee arthroplasty by health care region and period. | 31 |
| Table 9. Demographic and patient treatment profile of patients with primary and revision total knee arthroplasty by period..... | 33 |
| Table 10. Demographic and patient treatment profile of patients with primary and revision partial knee arthroplasty by period | 34 |
| Table 11. Number of primary knee arthroplasties per type of arthroplasty and period. | 36 |

| | |
|--|----|
| Table 13. Number of primary partial knee arthroplasties per fixation technique by time period..... | 37 |
| Table 14. Number of primary total knee arthroplasties per fixation technique by sex. | 38 |
| Table 15. Number of primary partial knee arthroplasties per fixation technique by sex | 38 |
| Table 16. Number of primary total knee arthroplasties per fixation technique by age group...38 | |
| Table 17. Number of primary partial knee arthroplasties per fixation technique by age group..... | 39 |
| Table 18. Number of primary total knee arthroplasties by fixation technique | 39 |
| Table 19. Number of primary partial knee arthroplasties by fixation technique | 39 |
| Table 20. Five most frequent models in primary knee arthroplasty per fixation type and period. Source RACat. | 41 |
| Table 21. Five most frequent models in primary knee arthroplasty per type of arthroplasty and period. Source RACat. | 42 |
| Table 22. Estimated cumulative revision rate after a primary total knee arthroplasty..... | 47 |
| Table 23. Estimated cumulative revision rate after a primary partial knee arthroplasty..... | 47 |
| Table 24. Estimated cumulative revision rate after a primary total knee arthroplasty by fixation technique | 48 |
| Table 25. Estimated cumulative revision rate after a primary partial knee arthroplasty by fixation technique | 49 |
| Table 26. Estimated cumulative revision rate after a primary total knee arthroplasty (CR/PS) .. | 50 |
| Table 27. Estimated cumulative revision rate after a primary partial knee arthroplasty (Unicompartmental/Patellofemoral) | 50 |
| Table 28. Number of hip arthroplasties by health region, period and by type of arthroplasty . | 53 |
| Table 29. Number of hip arthroplasties by health care region, centre, period and by type of arthroplasty. | 54 |
| Table 30. Number and percentages of classified primary hip arthroplasties reported by period | 57 |
| Table 31. Number of hip arthroplasties with the operated side (left or right) reported by period..... | 57 |
| Table 32. Age and sex of patients undergoing a primary total hip arthroplasty by health care region and period | 59 |

| | |
|--|----|
| Table 33. Age and sex of patients undergoing a primary partial hip arthroplasty by health care region and period..... | 59 |
| Table 34. Demographic and patient treatment profile of patients with primary and revision total hip arthroplasty by period. | 61 |
| Table 35. Demographic and patient treatment profile of patients with primary and revision partial hip arthroplasty by period. | 62 |
| Table 36. Number of primary hip arthroplasties per type of arthroplasty and period | 64 |
| Table 37. Number of primary total hip arthroplasties per fixation technique by time period ... | 65 |
| Table 38. Number of primary partial hip arthroplasties per fixation technique by time period | 65 |
| Table 39. Number of primary total hip arthroplasties per fixation technique and sex | 65 |
| Table 40. Number of primary partial hip arthroplasties per fixation technique and sex | 65 |
| Table 41. Number of primary total hip arthroplasties per fixation technique by age group..... | 65 |
| Table 42. Number of primary partial hip arthroplasties per fixation technique by age group.. | 67 |
| Table 43. Number of total conventional primary hip arthroplasties by type of friction device per age groups. | 67 |
| Table 44. Distribution of friction device per head size among conventional hip arthroplasties | 67 |
| Table 45. Five most frequent models in primary hip arthroplasty per type of arthroplasty and period..... | 69 |
| Table 46. Estimated cumulative revision rate after a primary total hip arthroplasty..... | 74 |
| Table 47. Estimated cumulative revision rate after a primary partial hip arthroplasty..... | 74 |
| Table 48. Estimated cumulative revision rate after a primary total hip arthroplasty by fixation technique | 75 |
| Table 49. Estimated cumulative revision rate after a primary partial hip arthroplasty by fixation technique | 76 |
| Table 50. Estimated cumulative revision rate after a primary total conventional hip arthroplasty per friction device | 77 |

LIST OF FIGURES

| | |
|---|----|
| Figure 1. Data review circuit | 19 |
| Figure 2. Volume of knee arthroplasties sent to the RACat per year, and by type of arthroplasty | 24 |

| | |
|---|----|
| Figure 3. Completeness of data sent to the RACat per year, and by type of arthroplasty. | 28 |
| Figure 4. Percentage of completeness of primary knee arthroplasties per centre during the period 2013-2014 | 28 |
| Figure 5. Percentages of classified primary knee arthroplasties per centre in relation to the volume of data sent in the period 2013-2014..... | 29 |
| Figure 6. Percentage of primary arthroplasties and knee replacements with reported laterality | 29 |
| Figure 7. Age of patients undergoing a primary knee arthroplasty by period | 32 |
| Figure 8. Percentage distribution of primary knee arthroplasties by age group. | 35 |
| Figure 9. Percentage distribution of the fixation technique in primary knee arthroplasties by time period | 37 |
| Figure 10. Number of different prosthesis models in relation to the volume of primary knee arthroplasties in the 2005-2014 period..... | 40 |
| Figure 11. Distribution of the follow-up time of the patients with primary knee arthroplasties | 45 |
| Figure 12. Estimated cumulative revision rate after a primary total knee arthroplasty by fixation technique | 49 |
| Figure 13. Cumulative incidence of revision after a primary total and partial knee arthroplasty adjusted by competitive risk of death, sex and age | 51 |
| Figure 14. Volume of hip arthroplasties sent to the RACat per year | 53 |
| Figure 15. Completeness of data sent to the RACat per year, and by type of arthroplasty | 56 |
| Figure 16. Percentage of completeness of primary hip arthroplasties per centre during the period 2013-2014 | 56 |
| Figure 17. Percentage of primary hip arthroplasties classified per centre according to the volume of data sent in the period 2013-2014..... | 57 |
| Figure 18. Percentage of primary and revision hip arthroplasty with operated side informed per centre, and in relation to data volume sent in the period 2013-2014..... | 58 |
| Figure 19. Age of patients undergoing a hip arthroplasty by period | 60 |
| Figure 20. Percentage distribution of primary hip arthroplasties by age group and type of arthroplasty | 64 |
| Figure 21. Percentage distribution of the fixation technique of primary hip arthroplasties by time period. A) Total arthroplasties; B) Partial arthroplasties | 66 |

| | |
|--|----|
| Figure 22. Number of different prostheses models in relation to the volume of primary hip arthroplasties for each centre sent to the RACat in the 2013-2014 period | 68 |
| Figure 23. Distribution of the follow-up time of patients undergoing a hip arthroplasty | 72 |
| Figure 24. Cumulative risk of revision after a total hip arthroplasty by fixation technique | 76 |
| Figure 25. Cumulative incidence of revision after a partial hip arthroplasty adjusted by competitive risk of death, age and sex..... | 78 |

ABSTRACT

Background and objectives

The Catalan Arthroplasty Register (RACat, for its acronym in Catalan) is a population-based register designed to assess the clinical effectiveness of hip and knee arthroplasties in Catalonia. The objectives of this report are 1) to present the results corresponding to the period 2005-2014, in terms of volume, completeness and quality of the information, as well as the temporary evolution of the arthroplasties and the variability per centre; 2) to describe the characteristics of patients undergoing a knee or hip arthroplasty, along with fixation techniques, models, and characteristics of the implants; and 3) to analyse the survival of knee and hip prosthesis in the short (at 1 year), mid- (at 5 years), and long-term (at 9 years).

Methodology

The RACat population comprises all insured citizens by Catalan Health Service (CatSalut) who underwent hip or knee arthroplasty in a healthcare centre run by the Integrated Healthcare System of Public Healthcare Network of Catalonia (SISCAT, for its acronym in Catalan). The current report covers the period between January 2005 and December 2014. During this period 53 of 61 centres participated in a voluntary basis, sending data to the RACat. By using a secure platform integrated in the Health Registers platform (RSA) of the CatSalut, hospitals sent data on patients, implants, interventions and surgical techniques. This information is complemented with data from the Central Register of Insured Persons (RCA, for its acronym in Catalan), which provides information on the patient's date of birth, place of residence, and vital status. Then, data is linked to the Minimum Basic Data Set at Hospital Discharge (MBDSHD) and the prosthesis catalogue. A descriptive analysis of the volume of arthroplasties included in the RACat per year and period (2005-2006; 2007-2008; 2009-2010; 2011-2012; 2013-2014), hospital, and healthcare region was carried out. The completeness of the RACat data was calculated per year and was compared against participating centres by means of a funnel plot. To assess the quality of the data we computed the percentage of primary knee and hip arthroplasties that were identified following the classification system used in the RACat and the percentage of cases sent to the RACat with the operated side informed. The characteristics of patients (sex, age, reason for surgery, comorbidities), the characteristics of the arthroplasties (type of arthroplasty, fixation technique, friction device) and the most frequently used models, were all described. The cumulative revision rate for knee and hip arthroplasties was estimated for partial and total procedures, by type of fixation and per type of arthroplasty, and by different time intervals: 0 to 30 days, 30 to 90 days, 90 days to 1 year, 1 to 3 years, 3 to 5 years, 5 to 7 years and 7 to 9 years. Cox proportional-hazard regression models were fitted to explore the effects of different factors in the revision arthroplasties adjusted by different covariates.

Results

Knee arthroplasties

- During 2005-2014, the RACat recorded a total of 60,192 primary knee arthroplasties and 6,689 revision knee arthroplasties. The highest completeness was reached in 2011,

showing a slight decrease over the latter period, coinciding with technical changes in data notification. In the latter period (2013-2014), the percentage of arthroplasties that were classified was 81.9% and the percentage of those with reported laterality available was 99.9%.

- In the time period spanning between 2013 and 2014, the mean age of patients undergoing a total and partial knee arthroplasty was 72.3 and 66.8 years, respectively. Osteoarthritis was the main reason for intervention both for total and partial procedures (98.7% and 96.9%, respectively) and mechanical complications were the main cause for revision both for total and partial knee arthroplasties.
- Total knee procedures represented 97.0% of knee arthroplasties, whereas partial knee arthroplasties accounted for 3.0%. Within total knee arthroplasties, the most common arthroplasties were the ones that preserve the posterior cruciate retaining (CR) and the posterior stabilized ones (PS), representing 46.0% and 51.7%, respectively. Among partial knee procedures, unicompartmental arthroplasties were the most common type (82.9%). A 75.7% and 81.5% of total and partial knee arthroplasties were cemented, with a similar distribution per periods, sex, and age groups.
- During 2005-2014, a total of 137 models (counting separately the cemented and cementless models) were identified. The number of models used by the participating hospitals ranged from 1 to 38, with an average of 12 models per hospital.
- The cumulative revision rate for total knee arthroplasties was 0.98%, 3.84% and 5.17% at 1-, 5- and 9-years follow-up, respectively, whereas the cumulative revision rate for partial knee arthroplasties was 2.07% at 1-year and 10.24% at 5-years follow-up. For total knee arthroplasties, an increased risk of revision for patients under 65 years and for patients aged between 65 and 75 years compared to those aged 85 years or over was observed (HR=3.34 (IC95%: 1.96-5.70); HR=2.02 (IC95%: 1.19-3.43), respectively).

Hip arthroplasties

- During 2005-2014, the RACat recorded a total of 46,488 primary hip arthroplasties and 5,237 revision hip arthroplasties. The highest completeness was reached in 2011, showing a slight decrease over the latter period. In the last period (2013-2014), the percentage of arthroplasties classified was 77.1% and the percentage of those with reported laterality was 97.6%.
- In the 2013-2014 time period, the mean age of patients undergoing a total and partial hip arthroplasty was 68.5 and 84.4 years, respectively. Osteoarthritis was the main reason for intervention for total procedures (81.7%), whereas fracture was the reason for intervention in 95.6% of partial procedures. Mechanical complications were the main cause for revision both for total and partial hip arthroplasties.
- Total hip procedures represented 77.6% of hip arthroplasties, whereas partial procedures accounted for 22.4%. Over 95% of total hip arthroplasties were total conventional implants. Within partial hip arthroplasties, bipolar and unipolar monoblok were the most common types, representing a 46.7% and 47.2%, respectively. The proportion on cementless arthroplasties was 64.4% and 50.5% in total and partial hip procedures, respectively. In conventional total hip arthroplasties, the most commonly used friction device was metal-polyethylene (65.8%). Small heads (<32mm) were commonly used among metal-polyethylene, ceramic-polyethylene, and metal-metal implants (72.7%, 68.1% and 43.3%, respectively). Medium heads (32-36mm) were mostly common among ceramic-ceramic arthroplasties (71%).
- During 2005-2014, 1,114 models of stem-socked combinations were identified. The number of acetabular models used by the hospitals ranged from 1 to 34, with an average of 16 per hospital. As for stem models, the number ranged from 1 to 33 with an average of 15 models per hospital.

- The cumulative revision rate for total hip arthroplasties was 1.29%, 2.99% and 4.57% at 1-, 5- and 9-years follow-up, respectively, whereas the cumulative revision rate for partial hip arthroplasties was 1.23%, 2.11% and 2.38%, respectively. For total hip arthroplasties, an increased risk of revision was observed for patients under 65 years, patients aged 65-75 years, and patients between 75-85 years compared to those aged 85 years or over (HR=2.87 (CI 95%: 1.47-5.61); HR=2.38 (CI 95%: 1.22-4.65); HR= 1.96 (CI 95%: 1.00-3.84)). A similar trend was observed for partial hip arthroplasties for patients aged 65-75 years and 75-85 years (HR=2.81 (CI 95%: 1.78-4.45; HR=1.63 (CI 95%: 1.21-2.20)). The Cox model considering the different friction devices and adjusted by age, sex, and comorbidities revealed an increased revision rate for metal-metal prosthesis in comparison to metal-polyethylene prosthesis (HR=1.43 CI 95%: 1.06-1.93).

Conclusions

The results presented in this report provide an overview of the quality of RACat's data, the characteristics of arthroplasties carried out in Catalonia over the last 10 years, and the clinical effectiveness of primary knee and hip arthroplasties up to 9 years of follow-up. Although the volume and quality of data recorded in the RACat presented a positive trend over time, there is significant variability per centre, which indicates room for improvement in the process of data acquisition and the need for defining strategies aimed at ensuring better quality reporting. In terms of clinical effectiveness, the results presented are consistent with the international literature. In our context, the RACat is being consolidated as a useful tool to analyse variations in patient characteristics and prosthesis utilization among hospitals or health regions in Catalonia, and may become a powerful system for post-market surveillance of implants. Improvements in data quality, the incorporation of new variables, the continuing increase of the follow-up period, along with consistently working closely with clinicians and key stakeholders, will enable to perform more robust and meaningful analysis that hopefully can foster improvements for the healthcare system.

PRESENTATION

The Catalan Arthroplasty Register (RACat, for its acronym in Catalan) was launched in 2005 in response to a common interest of the Catalan Health Service (CHS), the Catalan Society of Orthopaedic Surgery and Traumatology (SCCOT, for its acronym in Catalan), and the Agency for Health Quality and Assessment of Catalonia (AQuAS, for its acronym in Catalan). The RACat is a population-based register designed to assess the clinical effectiveness of hip and knee arthroplasties in Catalonia. Using the CHS Information System, publicly funded hospitals sent information to the RACat on knee and hip arthroplasties, including patient identification, hospital, joint (hip/knee), type (primary/revision), side of operation, date of surgery, and implants. The information is complemented by linking this data with other databases with the aim to complete the patient's information, clinical procedures, and patient follow-up. The RACat was made possible thanks to the cooperation of professionals and participant centres, along with the collaboration of manufacturers, which help with complementing implant information.

The 2005-2014 Hip and Knee Report is based on the analysis of 118,606 primary and revision knee and hip arthroplasties recorded by the Catalan Arthroplasty Register (RACat) with a procedure date up to and including 31 December 2014. It provides an update of the results presented before incorporating certain new aspects in comparison with previous reports. Firstly, the results for knee arthroplasties are presented separately for total and partial arthroplasties. This differentiation provides a more accurate characterization of the implant outcomes and patient characteristics. As had been the case in previous reports, the results for hip arthroplasties are also presented by total and partial arthroplasties. Secondly, giving the increase of the follow-up period, we are able to show survival outcomes up to 9 years of follow-up. Nevertheless, the number of cases with long term follow up is relatively small for some subgroups, so these figures must be interpreted with caution when reading the current report. Lastly, we have incorporated information on head size for total conventional hip arthroplasties. This data provides a more complete overview for the characterization of hip implants and represents an opportunity to generate further analyses.

1. INTRODUCTION

As has been the case in most European countries, increases in degenerative illnesses due to the population ageing and changes in indication criteria, has led to a significant rise in the number of hip and knee arthroplasties in Catalonia in recent years¹⁻⁴. Although these surgeries are among the most cost-effective medical procedures, significant variability in the clinical practice, along with waiting list problems and a high impact on healthcare budget have been highlighted⁵.

After 10 years of operation, the RACat has been consolidated as a useful tool for information, evaluation, and guidance in clinical practice in Catalonia, and has also provided relevant data for professionals, public health bodies and manufacturers. In the context of Spain, the RACat is a pioneering initiative that will be very valuable if a nationwide registry is eventually launched. On an international level, the RACat adds to more than 50 arthroplasty registers that are currently underway around the world, and collaborates with several networks on a European and global level.

This report aims to provide updated information on knee and hip arthroplasties carried out in publicly funded Catalan hospitals between 2005 and 2014. In the following pages, you will find data on quality and completeness of the information, patient's and implant characteristics, and prosthesis survival in the short, medium, and long-term.

2. OBJECTIVES

The objective of this report is to present the results of the information sent to the RACat during the period 2005-2014 regarding the characteristics of patients, arthroplasty interventions and implanted prostheses, along with the implant survival, and factors related with a better/worse prognosis of the primary hip and knee arthroplasties.

The specific objectives are:

1. To describe data sent to the RACat in terms of completeness and quality of the information, as well as the temporary evolution of the arthroplasties and the variability per centre.
2. To describe the characteristics of patients undergoing a knee or hip arthroplasty, the fixation techniques, models, characteristics of the implanted prostheses and trends over time.
3. To analyse the survival of knee and hip prosthesis in the short(at 1 year), medium (at 5 years), and long-term (at 9 years).

3. METHODOLOGY

3.1. Study population

The study population comprises all insured citizens by Catalan Health Service (CHS) who underwent hip or knee arthroplasty between January 2005 and December 2014 in a healthcare centre run by the Integrated Healthcare System of Public Healthcare Network of Catalonia (SIS-CAT, for its acronym in Catalan). During this period 53 of 61 centres participated in a voluntary basis, sending data to the RACat. They represented over 85% of the activity that took place in the public sector in Catalonia. So far, RACat does not routinely gather data from private centres, but some centres have begun to adapt their information systems to RACat requirements.

The information provided in this report concerns the characteristics of primary arthroplasties given that they represent the majority of interventions, versus revisions.

3.2. Data collection and sources of information

The RACat database is integrated in the Health Registers platform (RSA) of the CHS that includes a Web Service interface with other office applications, and is accessed through the Applications Portal of Health Department via on-line or Web Services. By using this platform, we ensure the secure transmission of data.

Hospitals sent patient data to the RACat on implants, interventions and surgical techniques. The quality of the information sent to the RACat is reviewed periodically. The results of the review process are sent to the members of the Technical Group of each centre to correct errors; the validated information is sent back to the RACat.

Data provided by centres are: a) patient's Personal Identification Code (PIC); b) date of admission; c) date of intervention; d) date of discharge; e) joint (knee or hip) f) whether it is a primary or a revision arthroplasty; g) operated side (right or left); h) name of the manufacturer; i) reference number and batch number of every prosthesis component informed.

Based on the patient identifier data, the register data are linked with other data sources (Table 1).

Central Insurance Register (RCA, for its acronym in Catalan). This is the register for persons covered by CHS, and provides information on the patient's date of birth, gender, place of residence, and status (death, alive or living outside Catalonia). Using the IIN sent by each centre, the information from the RCA is gathered with data sent by hospitals.

Minimum basic data set for hospital discharge (MBDSHD). This dataset provides information on the reason for intervention, the procedures performed and comorbidities. This dataset is provided by the CHS Activity Register Division.

Catalogue of prostheses. This is produced by the RACat and is based on information provided by manufacturers and distributors of prostheses. This catalogue made possible to complete the information sent by hospitals about implants (name of the manufacturer, reference number, batch number of all components implanted and the cement used), and to identify the type of

component, the fixation technique, the friction device (in total hip arthroplasties), whether fixed or mobile-bearing is used (in total knee arthroplasties), the prosthesis models implanted and to group the type of arthroplasty following the RACat's classification system (Table 2).

We are currently updating several variables included in the Catalogue related to prosthesis design, material, fixation, and component size. This information would allow us to carry out more detailed analysis and would enrich the current results. So far, the catalogue of prostheses offers information on more than 28,000 reference numbers pertaining to the primary components of hip prostheses (stem, acetabulum, insert and head) and knee (femoral, tibial, tibial insert, patella) of 67 manufacturers.

Table 1. Variables collected from RACat's source information

| Variables sent by hospital | Variables RCA | Variables MBDSHD | Catalogue of prostheses |
|---|---|--|---|
| <ul style="list-style-type: none"> • Personal Identification Code (PIC) • Centre • Clinical history number (CHN) • Date of admission and discharge* • Date of intervention • Type of arthroplasty (primary, revision) • Join (knee, hip) • Operated side (right, left) • Prosthesis information (manufacturer, reference and batch number) | <ul style="list-style-type: none"> • IIN • Sex • Date of birth • Status patient • Basic health unit • Health sector • Health care region | <ul style="list-style-type: none"> • IIN • Centre • CHN • Date of admission / discharge • Health care region • Sex • Principal / secondary diagnosis (reason for surgery; comorbidities) • Principal procedure • Hospital stay • Discharge from social and health services centres • Level of hospital complexity | <ul style="list-style-type: none"> • Manufacturer • Reference number • Trademark / brand name • Joint type • Type of implant • Design • Modularity • Mobility • Fixation • Fixation surface • Bearing surface (MoP, PoC, CoC, MoM) • Size • Side |

Data of discharge is available from November 2013 onwards

RCA: Central Insurance Register; MBDSHD: Minimum basic data set for hospital discharge; MoP: Metal on Polyethylene; PoC: Polyethylene on Ceramic; CoC, Ceramic on Ceramic; MoM: Metal on Metal.

RACat's surgical form. A surgical form was agreed by the Catalan Society of Orthopaedic Surgery and Traumatology (SCCOT, for its acronym in Catalan). This form includes information on the patient, surgery, intervention, implants and intraoperative complications⁶. Since the end of 2013, the variables of the surgical form have been integrated into RACat's information System. Some hospitals have started to report the variables of this form, but we are still in the implementation phase.


3.3. Data protection and confidentiality

The RACat is part of the CHS Information System, guaranteeing compliance with the applicable regulations on data protection (RD994/1999, of June 11; organic Law 15/1999, of 13 December; RP 428/1993, of 26 March) and ensuring that the information is transmitted via secure communication. In accordance with legal regulations in force, data is anonymised to carry out specific analysis.

3.4. Quality control and data validation

The quality of the information sent to the RACat is reviewed periodically. As part of the review process, we check the information related to prosthesis, patients, and arthroplasties. The data review circuit is detailed in Figure 1.

Figure 1. *Data review circuit*



3.5. Primary arthroplasty classification system

In order to group the different types of arthroplasties implanted, centres must send the necessary information regarding prosthesis components. RACat uses the classification system showed in Table 2, which was adapted from the Australian Orthopaedic Association National Joint Replacement⁷.

Primary knee arthroplasties

The knee joint is divided into three major compartments: patellofemoral, medial and lateral femorotibial joint. The type of arthroplasty is defined depending on the number of compartments replaced (Table 2): a) Partial knee arthroplasties: when one or two compartments are replaced (patellofemoral and/or femorotibial). b) total knee arthroplasty (TKR), when two or three compartments are replaced. Within partial knee arthroplasties we differentiate between: unicompartmental (UKA), when a single compartment is replaced (medial or lateral tibiofemoral); bicompartamental, when the medial or lateral and patellofemoral compartments are replaced;

and patellofemoral, if only this compartment is replaced. Furthermore, total arthroplasties are classified into different groups, depending on the joint range of motion and whether cruciate and collateral ligaments are preserved or not: cruciate retaining (CR), cruciate substituting or posterior stabilized (PS), constrained, hinged and tumoral (in these latter two categories collateral ligaments are not preserved).

Primary hip arthroplasties

Two large groups of arthroplasties have been defined: partial and total arthroplasty, when a part or all of the joint is replaced, respectively: a) partial hip arthroplasty (PHA), if femoral joint is replaced, and b) total hip arthroplasty (THA) when acetabulofemoral joint is replaced (Table 2). In addition, partial arthroplasties are grouped as: unipolar monoblock when only a stem is implanted, unipolar modular, when a stem and a cephalic component is implanted, and bipolar, where a stem and a bipolar head component are implanted. In the case of total arthroplasties, they are grouped as: conventional, when a stem, head and acetabulum are implanted; resurfacing, when a large head and cup are implanted and the femoral head is not removed; Dual mobility acetabular component (also known as unconstrained tripolar implant) when a dual-articulation acetabular cup is used; and with short stem, when a short stem, head and acetabulum are implanted.

Table 2. Classification system for primary hip and knee arthroplasties

| Type of arthroplasty | |
|---|---|
| Knee | Hip |
| Partial knee arthroplasty (PKA) Patellofemoral Unicompartmental femorotibial (UKA) Bicompartamental Total knee arthroplasty (TKA) Cruciate retaining (CR) Posterior stabilised (PS) Constrained Hinged Tumoral | Partial hip arthroplasty (PHA) Unipolar monoblock Unipolar modular Bipolar Total hip arthroplasty (THA) Resurfacing With short stem Conventional Dual Mobility o tripolar |

3.6. Statistical analysis

The analyses were structured in three areas, in line with the proposed objectives: quality and completeness of the data, descriptive analyses and survival analyses.

Completeness and quality of the information

To analyse the quality of the data and the completeness of the information, the information from the RACat and the Minimum Basic Data Set at Hospital Discharge (MBDSHD) was considered. To this end, the analysis accounted for all patients covered by CHS who received a knee or hip arthroplasty in one of the centres collaborating with the RACat (n=118,606) from January 2005 to December 2014, including 66,881 knee arthroplasties and 51,725 hip arthroplasties. We only considered the public activity carried out in Catalonia over the study period.

MBDSHD is a comprehensive clinical and administrative database on all health and morbidity activity conducted in public and private hospitals in Catalonia. Thus, the activity of the 53 hospitals participating in the RACat is recorded herein. We used this database as a gold standard to calculate the completeness of the RACat. The information of both registers was linked together using the patient's IIN, centre, admission date, joint, and type of arthroplasty (primary or revision).

A descriptive analysis of the number of arthroplasties (**volume**) included in the RACat per year and period (2005-2006; 2007-2008; 2009-2010; 2011-2012; 2013-2014) was carried out. The completeness of the RACat's data, per centre and year, was calculated. **Completeness** was considered to be the ratio between the arthroplasties volume sent to the RACat and the arthroplasties volume sent to the MBDSHD by participating hospitals during the 2005-2014 period. A funnel plot between the completeness of each centre and the case volume sent to the RACat, in an anonymized way, was plotted. To do this, the overall completeness of all participating centres of the RACat with a 95 % confidence interval was used as a reference.

Two indicators were defined to assess the **quality of the data**: the percentage of primary knee and hip arthroplasties that were identified following the classification system used in the RACat (Table 2) and the percentage of cases sent to the RACat with the operated side informed (right or left). A funnel plot for each indicator with the result of every centre and the case volume sent to the RACat, was also anonymous. It was used as a reference for the overall result with a 95 % confidence interval.

The funnel plot is recommended as a visual aid to benchmark centres. In this type of graph, an estimate of an indicator versus an interpretable precision measure is drawn. The limits of control draw the shape of a funnel over the objective result control. In the case of centres that are above or below the limits of control, the value of the indicator will be interpreted as being above or below the set or target outcome.

Descriptive analyses of patients and arthroplasties

To perform the descriptive analyses, all cases in which there was correspondence between the data from the RACat and the MBDSHD were selected, thus ruling out 6.8% of cases. This meant a total selection of 110,618 arthroplasties (61,565 knee arthroplasties and 47,428 hip arthroplasties).

For each type of joint (hip and knee), a descriptive analysis was carried out (first by overall data and then by health care regions) of the patient characteristics (volume, sex, age and type of arthroplasty). At the same time, and only for those patients included in the RACat and linked with the MBDSHD, other questions related to morbidity and the care delivery process were described (reason for surgery, comorbidities, hospital stay and percentage of referrals to a convalescence centre at hospital discharge).

The **global burden of revision**, the percentage of revisions sent to the RACat compared to total number of interventions sent (primary and revisions), was calculated. This index provides the percentage of revisions undertaken in a given period in relation to the total number of interventions. In contrast with the cumulative revision rate, revision procedures described in the burden revision during 2005-2014, do not necessarily correspond to patients who underwent a primary arthroplasty in the same period (it is a cross-sectional cut without a follow-up).

Fixation technique (cemented, uncemented and hybrids) and the type of arthroplasty used were described per age group and period; and, in the case of conventional hip arthroplasties, the **friction device** (metal-metal, metal-polyethylene, ceramic-polyethylene and ceramic-ceramic) was also described overall and per age group. The five most common **implant models** have been described by type of arthroplasty and fixation technique.

Information on **head size** among conventional hip arthroplasties was also analysed. This variable was grouped in 3 categories: small heads, if the head diameter was <32mm; medium heads, when the head diameter ranged between 32 and 36 mm; and large heads, when the head diameter was >36mm.

Implant survival analysis

As for the survival analyses, these excluded not only cases where there was no data link between the RACat and the MBDSHD, but also cases where the side (right or left) could not be determined, and cases in which information was received about the revision surgery but not about the primary surgery. As a result, these analyses covered a total of 46,406 knee arthroplasties and 31,603 hip arthroplasties.

A survival analysis was conducted considering the time elapsed from the primary arthroplasty until the revision of the prosthesis, irrespective of the reason for revision. A revision arthroplasty was defined as any procedure involving removal, exchange or addition of any implant part.

Cases were censored at the date of death, at the date of discharge from Catalonia to other Spanish autonomous communities or foreign countries (patients no longer affiliated to the CHS) or at 31st December 2014 if no event occurred before the end of the study, whatever occurred first.

The **cumulative revision rate** for knee arthroplasties was estimated for partial and total procedures, by the type of fixation and according to the type of arthroplasty. For hip arthroplasties, differentiated analyses were carried out for partial and total procedures. For the latter ones, the cumulative rate by friction device was also considered. When calculating the cumulative revision rates, the actuarial adjustment by the number of patients at risk of replacement was considered for each time interval.

To find out the revision rates at short-term or to year basis, the following intervals of time were used: 0 to 30 days, 30 to 90 days, 90 days to 1 year, 1 to 3 years, 3 to 5 years, 5 to 7 years and 7 to 9 years.

On the other hand, Cox proportional-hazard regression models were fitted to explore the effects of different factors in the revision arthroplasties adjusted by different covariates⁸. The hazard ratio (HR) of revision was estimated depending on the type of fixation, adjusted by patient age and sex. These models allow considering the elapsed time until the event or the censorship, with proportional risk assumption. The reading of a HR corresponds to that of a relative risk. Therefore, a HR>1, where 1 is not included in the 95% confidence intervals (CI), measures how high the risk for replacement in patients with the factor is, in relation to the ones that do not have it.

In order to adjust the analyses by risk of patient death, we also fitted Cox proportional-hazard models for competing risks of Fine and Grey⁹.

4. KNEE ARTHROPLASTIES RESULTS

4.1. Scope and quality of data

In the period described, data from 60,192 patients with primary arthroplasty and from 6,689 patients with revision knee arthroplasty were received. As shown in Figure 2 and Table 3 the volume of data increased over the first three periods in most health regions, suffering a slight decrease from 2012 onwards.

Table 4 shows the volume of primary and revision arthroplasties sent during the study period, per centre. The overall burden of revisions, understood as the percentage of revision arthroplasties compared to the total number of events sent, was 10.0%. Remarkable disparities were found between centres with a minimum burden of 0% and a maximum of 24.8%.

Figure 3 shows the completeness by type of knee arthroplasty (primary or revision) over the years. It can be observed that the highest completeness both for primaries and revisions was reached in 2011. A gradual decrease was observed from 2012 onwards, coinciding with technical changes in data notification. Figure 4 shows the variability in data completeness of each centre (blue points) compared with overall completeness (green line) and a confidence interval based on overall completeness. The centres that are below the range of the confidence interval exhibit lower completeness compared to the overall figure, and the ones that are above exhibit higher completeness than the overall. Only 9 out of the 53 centres fell below the confidence interval based on the overall reference value.

As a quality indicator of data, we calculated the percentage of arthroplasties that had been classified following the RACat's classification. Considering the whole period, the percentage of classified primary arthroplasties was 82.2% (Table 5). Figure 5 is a funnel plot showing the percentage of classified primary arthroplasties per centre in relation to the volume of data sent in the period 2013-2014. As can be seen, almost all the healthcare centres present percentages around 80%, which indicates good data quality of the information sent.

The percentage of arthroplasties with reported laterality increased over the study period reaching 99.9% in 2013-2014 (Table 6). Considering data from the latter 2 years we plotted percentage of laterality reported per centre (Figure 6). Here we can see that all centres reported correctly this variable correctly in 2013-2014.

Figure 2. Volume of knee arthroplasties sent to the RACat per year, and by type of arthroplasty. Source RACat




Table 3. Number of knee arthroplasties by health region period and by type of arthroplasty. Source RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|----------------------------|--------------|------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| Lleida | 216 | 15 | 515 | 31 | 604 | 48 | 640 | 27 | 798 | 8 | 2,773 | 129 |
| Camp de Tarragona | 129 | 3 | 817 | 30 | 1,157 | 46 | 1,014 | 31 | 1,017 | 65 | 4,134 | 175 |
| Terres de l'Ebre | 0 | 0 | 0 | 0 | 637 | 59 | 400 | 36 | 305 | 48 | 1,342 | 143 |
| Girona | 1,190 | 83 | 1,511 | 118 | 1,829 | 169 | 1,496 | 159 | 1,687 | 102 | 7,713 | 631 |
| Catalunya central | 615 | 43 | 1,137 | 93 | 1,218 | 176 | 1,164 | 200 | 1,282 | 153 | 5,416 | 665 |
| Alt Pirineu i Aragó | 97 | 3 | 116 | 0 | 135 | 3 | 134 | 6 | 124 | 0 | 606 | 12 |
| Barcelona | 5,576 | 476 | 7,491 | 773 | 9,239 | 1,189 | 8,317 | 1,079 | 7,585 | 1,417 | 38,208 | 4,934 |
| Total | 7,823 | 623 | 11,587 | 1,045 | 14,819 | 1,690 | 13,165 | 1,538 | 12,798 | 1,793 | 60,192 | 6,689 |

Table 4. Number of knee arthroplasties by health care region, centre, period and by type of arthroplasty. Source RACat.

| Hospital | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|---|-----------|----------|-----------|----------|-----------|----------|-----------|----------|-----------|----------|--------------|-------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | % revision |
| Lleida | | | | | | | | | | | | |
| Hospital Universitari Arnau de Vilanova | 0 | 0 | 0 | 2 | 108 | 12 | 161 | 4 | 255 | 2 | 524 | 3.7 |
| Hospital de Santa Maria | 164 | 9 | 307 | 20 | 324 | 21 | 284 | 18 | 330 | 4 | 1,409 | 4.9 |
| Clínica de Ponent | 52 | 6 | 208 | 9 | 172 | 15 | 195 | 5 | 213 | 2 | 840 | 4.2 |
| Camp de Tarragona | | | | | | | | | | | | |
| Hospital Universitari de Tarragona Joan XXIII | 1 | 0 | 129 | 17 | 254 | 23 | 277 | 12 | 323 | 24 | 984 | 7.2 |
| Hospital Universitari Sant Joan de Reus | 2 | 0 | 92 | 0 | 313 | 0 | 241 | 1 | 127 | 0 | 775 | 0.1 |
| Hospital de Sant Pau i Santa Tecla | 36 | 0 | 203 | 6 | 220 | 13 | 211 | 10 | 237 | 12 | 907 | 4.3 |
| Pius Hospital de Valls | 87 | 3 | 156 | 3 | 120 | 4 | 112 | 2 | 132 | 10 | 607 | 3.5 |
| Hospital del Vendrell | 3 | 0 | 237 | 4 | 250 | 6 | 173 | 6 | 198 | 19 | 861 | 3.9 |
| Terres de l'Ebre | | | | | | | | | | | | |
| Hospital Tortosa Verge de la Cinta | 0 | 0 | 0 | 0 | 309 | 59 | 200 | 36 | 227 | 48 | 736 | 16.3 |
| Hospital Comarcal Móra d'Ebre | 0 | 0 | 0 | 0 | 328 | 0 | 200 | 0 | 78 | 0 | 606 | 0.0 |
| Girona | | | | | | | | | | | | |
| Hospital Universitari de Girona Josep Trueta | 0 | 0 | 243 | 34 | 173 | 53 | 243 | 47 | 389 | 12 | 1,048 | 12.2 |
| Clínica Girona SA | 0 | 0 | 0 | 0 | 200 | 11 | 167 | 4 | 122 | 0 | 489 | 3.0 |
| Clínica Salus Infirmorum | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 39 | 3 | 39 | 7.1 |
| Hospital Comarcal de Blanes | 287 | 35 | 358 | 29 | 384 | 26 | 266 | 20 | 272 | 14 | 1,567 | 7.3 |
| Hospital de Campdevanòl | 46 | 0 | 51 | 0 | 53 | 0 | 44 | 0 | 63 | 0 | 257 | 0.0 |
| Hospital de Figueres | 178 | 13 | 343 | 26 | 356 | 28 | 257 | 37 | 91 | 8 | 1,225 | 8.4 |
| Hospital de Palamós | 272 | 20 | 282 | 12 | 277 | 18 | 191 | 19 | 234 | 38 | 1,256 | 7.9 |
| Hospital de Sant Jaume d'Olot | 223 | 13 | 44 | 7 | 149 | 17 | 86 | 18 | 145 | 12 | 647 | 9.4 |
| Hospital Provincial Santa Caterina | 184 | 2 | 190 | 10 | 237 | 16 | 242 | 14 | 332 | 15 | 1,185 | 4.6 |
| Catalunya Central | | | | | | | | | | | | |
| Centre Hospitalari-ALTHAIA | 372 | 37 | 778 | 62 | 674 | 127 | 610 | 134 | 598 | 103 | 3,032 | 13.3 |
| Hospital General de Vic | 154 | 1 | 123 | 4 | 224 | 7 | 288 | 33 | 372 | 32 | 1,161 | 6.2 |
| Hospital de Sant Bernabé | 0 | 0 | 51 | 4 | 122 | 8 | 103 | 8 | 150 | 5 | 426 | 5.5 |
| Fundació Sanitària d'Igualada FP | 89 | 5 | 185 | 23 | 198 | 34 | 163 | 25 | 162 | 13 | 797 | 11.2 |

| Hospital | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|--|-----------|----------|-----------|----------|-----------|----------|-----------|----------|-----------|----------|--------------|-------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | % revision |
| Alt Pirineu i Aran | | | | | | | | | | | | |
| Fundació Sant Hospital de la Seu d'Urgell | 50 | 1 | 46 | 0 | 51 | 3 | 38 | 4 | 46 | 0 | 231 | 3.4 |
| Hospital Comarcal del Pallars | 22 | 0 | 29 | 0 | 64 | 0 | 64 | 0 | 65 | 0 | 264 | 0.0 |
| Hospital de Puigcerdà | 25 | 2 | 21 | 0 | 20 | 0 | 32 | 2 | 13 | 0 | 111 | 3.5 |
| Barcelona | | | | | | | | | | | | |
| Hospital de Viladecans | 0 | 0 | 0 | 0 | 108 | 13 | 190 | 14 | 177 | 1 | 475 | 5.6 |
| Hospital Universitari de Bellvitge | 292 | 0 | 276 | 0 | 482 | 11 | 351 | 16 | 200 | 4 | 1,601 | 1.9 |
| H.Universitari Germans Trias i Pujol | 0 | 0 | 35 | 3 | 376 | 36 | 346 | 18 | 463 | 3 | 1,220 | 4.7 |
| Hospital Clinic i Provincial de Barcelona | 1,612 | 174 | 1,062 | 275 | 1,242 | 319 | 834 | 294 | 824 | 272 | 5,574 | 19.3 |
| Hospital de L'Hospitalet - CSI | 599 | 37 | 458 | 29 | 537 | 55 | 411 | 52 | 7 | 489 | 2,012 | 24.8 |
| Hospital Dos de Maig de Barcelona - CSI | 114 | 6 | 114 | 9 | 231 | 11 | 244 | 9 | 137 | 118 | 840 | 15.4 |
| Hospital Mútua de Terrassa | 0 | 0 | 389 | 44 | 581 | 52 | 493 | 68 | 158 | 23 | 1,621 | 10.3 |
| Hospital de l'Esperit Sant | 185 | 11 | 306 | 27 | 391 | 50 | 284 | 22 | 375 | 9 | 1,541 | 7.2 |
| Corporació Sanitària Parc Taulí | 231 | 32 | 639 | 84 | 630 | 82 | 542 | 63 | 326 | 96 | 2,368 | 13.1 |
| Hospital de Sant Boi- Parc Sanitari St Joan de Deu | 225 | 14 | 307 | 14 | 237 | 32 | 219 | 25 | 318 | 24 | 1,306 | 7.7 |
| Hospital de Sant Celoni - Fundació privada | 99 | 11 | 174 | 13 | 130 | 25 | 154 | 12 | 134 | 6 | 691 | 8.8 |
| Hospital de Terrassa | 0 | 0 | 257 | 28 | 287 | 27 | 239 | 17 | 371 | 16 | 1,154 | 7.1 |
| Parc de Salut del Mar | 139 | 27 | 414 | 26 | 506 | 88 | 750 | 125 | 0 | 0 | 1,395 | 14.7 |
| Hospital General de Granollers | 257 | 35 | 339 | 18 | 83 | 11 | 327 | 34 | 391 | 49 | 1,472 | 9.5 |
| Hospital Municipal de Badalona | 330 | 20 | 268 | 83 | 289 | 20 | 191 | 21 | 252 | 18 | 1,401 | 6.5 |
| Hospital Residència Sant Camil | 364 | 41 | 289 | 18 | 213 | 68 | 126 | 24 | 192 | 25 | 1,163 | 17.2 |
| Hospital Sant Joan de Déu de Martorell | 108 | 7 | 0 | 0 | 270 | 29 | 149 | 49 | 80 | 28 | 896 | 12.8 |
| Hospital de Sant Joan de D'Esplugues Llob | 0 | 0 | 0 | 0 | 1 | 0 | 0 | 0 | 1 | 0 | 2 | 0.0 |
| Hospital Sant Rafael | 0 | 0 | 0 | 0 | 378 | 54 | 302 | 57 | 368 | 49 | 1,048 | 13.3 |
| F.G.S. Hospital de la Santa Creu i Sant Pau | 0 | 0 | 0 | 0 | 226 | 29 | 394 | 50 | 427 | 54 | 1,047 | 11.3 |

| Hospital | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|-------------------------------------|--------------|------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|-------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | % revision |
| Barcelona (cont.) | | | | | | | | | | | | |
| H. de Plató | 0 | 0 | 133 | 7 | 150 | 10 | 58 | 2 | 77 | 3 | 418 | 5.0 |
| Hospital Universitari Sagrat Cor | 315 | 38 | 495 | 43 | 543 | 49 | 618 | 47 | 755 | 55 | 2,726 | 7.8 |
| Fundació Privada Hospital de Mollet | 135 | 7 | 146 | 11 | 152 | 4 | 220 | 6 | 289 | 1 | 942 | 3.0 |
| Hospital Comarcal de l'Alt Penedes | 0 | 0 | 166 | 13 | 165 | 29 | 136 | 12 | 186 | 22 | 653 | 10.4 |
| Hospital de Mataro | 116 | 16 | 412 | 28 | 448 | 30 | 275 | 17 | 280 | 26 | 1,531 | 7.1 |
| H. Universitari Vall d'Hebron | 455 | 0 | 812 | 0 | 583 | 55 | 464 | 25 | 797 | 26 | 3,111 | 3.3 |
| Total | 7,823 | 623 | 11,587 | 1,045 | 14,819 | 1,690 | 13,165 | 1,538 | 12,798 | 1,793 | 60,192 | 10.0 |

Figure 3. Completeness of data sent to the RACat per year, and by type of arthroplasty. Source RACat-MBD-SHD




Figure 4. Percentage of completeness of primary knee arthroplasties per centre during the period 2013-2014. Source: RACat




Table 5. Percentages of classified primary knee arthroplasties in relation to the volume sent to the RACat. Source: RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|-----------------------|--------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Classified | 4,637 | 59.3 | 9,409 | 81.2 | 12,925 | 87.2 | 12,003 | 91.2 | 10,486 | 81.9 | 49,460 | 82.2 |
| Not Classified | 3,186 | 40.7 | 2,178 | 18.8 | 1,894 | 12.8 | 1,162 | 8.8 | 2,312 | 18.1 | 10,732 | 17.8 |
| Total | 7,823 | 100.0 | 11,587 | 100.0 | 14,819 | 100.0 | 13,165 | 100.0 | 12,798 | 100.0 | 60,192 | 100.0 |

Figure 5. Percentages of classified primary knee arthroplasties per centre in relation to the volume of data sent in the period 2013-2014. Source: RACat.





Table 6. Number of knee arthroplasties with the operated side (left or right) reported by period. Source RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|---------------------|--------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Reported | 6,030 | 71.4 | 11,892 | 94.1 | 16,144 | 97.8 | 14,581 | 99.2 | 14,574 | 99.9 | 63,221 | 94.5 |
| Not reported | 2,416 | 28.6 | 740 | 5.9 | 365 | 2.2 | 122 | 0.8 | 17 | 0.1 | 3,660 | 5.5 |
| Total | 8,446 | 100.0 | 12,632 | 100.0 | 16,509 | 100.0 | 14,703 | 100.0 | 14,591 | 100.0 | 66,881 | 100.0 |

Figure 6. Percentage of primary arthroplasties and knee replacements with reported laterality per centre in relation to volume of data sent in the latter period. Source RACat.



4.2. Demographic and patient treatment process profile

From this section on, the results are presented split by total and partial knee arthroplasties.

The average age of patients undergoing a primary total knee arthroplasty was 72.2 years (standard deviation; SD: 7.6) and for those who underwent a partial knee arthroplasty was 65.5 years (SD: 9.7), being the percentage of women 71.4% and 67.0%, respectively (Table 7, Table 8).

In the time period spanning between 2005 and 2014, the age group distribution among patients with a total knee arthroplasty remained relatively stable (Figure 7). However, for those patients who experienced a partial knee arthroplasty the percentage of patients aged between 75 to 84 years increased over time.

Table 9 and Table 10 shows the demographic characteristics and treatment profile of patients who underwent total and partial knee arthroplasty, respectively. Patients undergoing a primary total knee arthroplasties were older than those undergoing a partial procedure, the mean age for the period 2013-2014 being 72.3 years and 66.8 years, respectively. The data shows that the main reason for intervention of arthroplasties carried out during the latter study period was osteoarthritis in 98.7% of primary total knee arthroplasties and 96.9% for partial knee arthroplasties. For the latter group, a gradual increase of this cause of intervention over time was observed. Mechanical complications were the main cause for revision both for total and partial knee arthroplasties. As expected by differences in patient's age distribution, the percentage of patients with 1 or more comorbidities was higher for patients undergoing total knee arthroplasties in comparison to those undergoing a partial knee arthroplasty.

Table 7. Age and sex of patients undergoing a primary total knee arthroplasty by health care regions and period. Source RACat.

| | 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | | Total | | |
|---------------------------|--------------|-------------------|-------------|--------------|-------------------|-------------|---------------|-------------------|-------------|---------------|-------------------|-------------|---------------|-------------------|-------------|---------------|-------------------|-------------|
| | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women |
| Lleida | 197 | 72.9 (6.2) | 68.0 | 460 | 73.9 (6.5) | 67.6 | 556 | 73.0 (6.8) | 66.7 | 562 | 73.3 (7.1) | 65.8 | 511 | 72.9 (7.5) | 68.1 | 2,286 | 73.3 (7.0) | 67.1 |
| Camp de Tarragona | 126 | 72.6 (6.0) | 71.4 | 643 | 72.4 (6.9) | 69.1 | 827 | 72.2 (6.8) | 71.5 | 756 | 71.8 (7.2) | 71.0 | 875 | 72.1 (7.2) | 67.5 | 3,227 | 72.2 (7.0) | 69.8 |
| Terres de l'Ebre | 0 | 0 | 0.0 | 0 | 0 | 0.0 | 301 | 71.4 (7.2) | 57.8 | 180 | 71.3 (7.7) | 61.1 | 161 | 71.6 (7.9) | 57.8 | 642 | 72.1 (7.6) | 58.7 |
| Girona | 926 | 72.2 (7.1) | 69.2 | 1,081 | 71.8 (7.2) | 68.2 | 1,314 | 72.1 (7.6) | 67.0 | 1,183 | 71.9 (8.2) | 67.3 | 1,445 | 72.0 (8.0) | 65.5 | 5,949 | 72.0 (7.7) | 67.2 |
| Catalunya central | 441 | 72.1 (7.4) | 72.8 | 1,029 | 71.6 (7.8) | 70.1 | 1,142 | 72.0 (7.5) | 68.9 | 1,107 | 72.1 (7.6) | 70.0 | 1,230 | 72.5 (7.5) | 70.8 | 4,949 | 72.1 (7.6) | 70.2 |
| Alt pirineu i Aran | 83 | 72.9 (8.0) | 62.7 | 114 | 73.4 (7.5) | 68.4 | 129 | 75.0 (7.7) | 64.3 | 128 | 73.6 (6.3) | 61.7 | 117 | 73.4 (7.0) | 65.0 | 571 | 73.7 (7.3) | 64.5 |
| Barcelona | 2,759 | 72.1 (7.3) | 74.2 | 5,855 | 72.2 (7.5) | 74.3 | 8,266 | 72.2 (7.7) | 73.5 | 7,738 | 72.3 (7.7) | 72.7 | 5,710 | 72.4 (7.8) | 72.0 | 30,328 | 73.2 (7.7) | 73.2 |
| Total | 4,532 | 72.2 (7.2) | 72.5 | 9,182 | 72.2 (7.4) | 72.3 | 12,535 | 72.2 (7.6) | 71.5 | 11,654 | 72.3 (7.7) | 71.1 | 10,049 | 72.3 (7.7) | 70.0 | 47,952 | 72.2 (7.6) | 71.4 |

Table 8. Age and sex of patients undergoing a primary partial knee arthroplasty by health care region and period. Source RACat.

| | 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | | Total | | |
|---------------------------|------------|-------------------|-------------|------------|-------------------|-------------|------------|-------------------|-------------|------------|--------------------|-------------|------------|-------------------|-------------|--------------|--------------------|-------------|
| | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women |
| Lleida | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 | 53.4 | 0 | 1 | 53.4 | 0 |
| Camp de Tarragona | 1 | 56.5 | 100.0 | 11 | 62.5 (7.4) | 81.8 | 8 | 66.4 (7.7) | 62.5 | 8 | 64.9 (8.0) | 62.5 | 6 | 60.4 (2.1) | 33.3 | 34 | 63.4 (7.1) | 64.7 |
| Terres de l'Ebre | 0 | 0 | 0.0 | 0 | 0 | 0.0 | 5 | 62.1 (4.6) | 60.0 | 1 | 61.2 | 100.0 | 0 | 0 | 0.0 | 6 | 61.9 (4.1) | 66.7 |
| Girona | 6 | 62.1 (8.1) | 83.3 | 15 | 65.1 (8.3) | 66.7 | 17 | 65.5 (7.9) | 64.7 | 8 | 58.5 (8.7) | 75.0 | 6 | 59.1 (4.1) | 66.7 | 52 | 63.7 (8.4) | 69.2 |
| Catalunya central | 5 | 53.8 (8.6) | 60.0 | 24 | 59.2 (9.0) | 79.2 | 37 | 57.6 (9.4) | 73.0 | 31 | 58.6 (8.8) | 58.1 | 25 | 59.5 (10.5) | 64.0 | 122 | 58.3 (9.3) | 68.0 |
| Alt pirineu i Aran | 6 | 72.4 (8.1) | 50.0 | 2 | 50.4 (1.1) | 50.0 | 6 | 63.9 (10.4) | 66.7 | 6 | 64.3 (11.0) | 83.3 | 1 | 65.3 | 100.0 | 21 | 64.8 (10.3) | 66.7 |
| Barcelona | 86 | 65.6 (8.3) | 67.4 | 171 | 65.2 (9.6) | 64.3 | 309 | 65.8 (9.4) | 68.0 | 287 | 66.5 (10.1) | 73.2 | 386 | 67.6 (9.2) | 62.4 | 1,239 | 66.4 (9.5) | 66.9 |
| Total | 104 | 65.2 (8.8) | 67.3 | 223 | 64.3 (9.5) | 66.8 | 382 | 65.0 (9.6) | 68.1 | 341 | 65.5 (10.0) | 71.9 | 425 | 66.8 (9.5) | 62.1 | 1,475 | 65.6 (9.7) | 67.0 |

Figure 7. Age of patients undergoing a primary knee arthroplasty by period. A) Total arthroplasties. B) Partial arthroplasties. Source RACat.





Table 9. Demographic and patient treatment profile of patients with primary and revision total knee arthroplasty by period. Source RACat-MBDSHD.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | |
|---------------------------------|-----------------|---------------|-----------------|---------------|------------------|---------------|------------------|---------------|-----------------|---------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n | n |
| Number | 4217 | 183 | 8,856 | 452 | 12,107 | 796 | 11,432 | 686 | 9,769 | 484 |
| % of Women | 3,068 (72.8) | 120 (65.6) | 6,417 (72.5) | 322 (71.2) | 8,663 (71.6) | 601 (75.5) | 8,133 (71.1) | 506 (73.8) | 6,831 (69.9) | 371 (76.7) |
| Mean age (SD) | 72.1 (7.1) | 73.1 (7.7) | 72.2 (7.4) | 72.6 (8.2) | 72.2 (7.6) | 73.0 (8.1) | 72.2 (7.7) | 73.3 (8.9) | 72.3 (7.7) | 73.0 (8.9) |
| Cause of intervention | | | | | | | | | | |
| Osteoarthritis | 4,111 (97.5) | | 8,681 (98.0) | | 11,905 (98.3) | | 11,264 (98.5) | | 9,646 (98.7) | |
| Other diagnostics | 106 (2.5) | | 175 (2.1) | | 202 (1.6) | | 168 (1.5) | | 123 (1.3) | |
| Cause of revision | | | | | | | | | | |
| Mechanical complications | | 119 (65.0) | | 253 (56.0) | | 518 (65.1) | | 485 (70.7) | | 357 (73.8) |
| Infection | | 26 (14.2) | | 34 (7.5) | | 81 (10.2) | | 53 (7.7) | | 20 (4.13) |
| Other diagnostics | | 38 (20.8) | | 165 (36.5) | | 197 (24.8) | | 148 (21.6) | | 107 (22.1) |
| Comorbidity | | | | | | | | | | |
| 1 or more | 2,459 (58.3) | 99 (54.1) | 5,667 (64.0) | 281 (62.1) | 8,206 (67.8) | 547 (68.7) | 8,224 (71.9) | 496 (72.3) | 7,229 (74.0) | 372 (76.8) |
| Healthcare discharge | 229 (5.43) | 8 (4.4) | 582 (6.6) | 39 (8.6) | 903 (7.5) | 59 (7.4) | 1185 (10.4) | 69 (10.1) | 870 (8.9) | 48 (9.9) |

Table 10. Demographic and patient treatment profile of patients with primary and revision partial knee arthroplasty by period. Source RACat-MBDSHD

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | |
|--|---------------|----------|---------------|----------|---------------|---------------|----------------|---------------|---------------|----------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) |
| Number | 94 | 0 | 219 | 0 | 371 | 3 | 336 | 2 | 424 | 0 |
| % of Women | 63 (67.0) | 0 | 145 (66.2) | 0 | 252 (67.9) | 1 (33.3) | 243 (72.3) | 2 (100) | 264 (62.3) | 0 |
| Mean age (SD) | 65.2 (8.9) | 0 | 64.4 (9.5) | 0 | 65.0 (9.6) | 68.4 (2.0) | 65.7 (10.2) | 65.7 (0.6) | 66.8 (9.5) | 0 |
| Cause of intervention | | | | | | | | | | |
| Osteoarthritis | 73 (77.7) | | 185 (84.5) | | 325 (87.6) | | 308 (91.7) | | 411 (96.9) | |
| Other bone and cartilage diseases | 14 (14.9) | | 18 (8.2) | | 21 (5.7) | | 18 (5.4) | | 5 (1.2) | |
| Other diagnostics | 7 (7.5) | | 16 (7.3) | | 25 (6.7) | | 10 (3.0) | | 8 (1.9) | |
| Cause of revision | | | | | | | | | | |
| Mechanical complications | | 0 | | 0 | | 2 (66.7) | | 1 (50.0) | | 0 |
| Other diagnostics | | 0 | | 0 | | 1 (33.3) | | 1 (50.0) | | 0 |
| Comorbidity | | | | | | | | | | |
| 1 or more | 34 (36.2) | 0 | 111 (50.7) | 0 | 216 (58.2) | 2 (66.7) | 214 (63.7) | 1 (50.0) | 261 (61.6) | 0 |
| Healthcare discharge | 0 | 0 | 0 | 0 | 1 (0.3) | 0 | 11 (3.3) | 0 | 7 (1.7) | 0 |

Revision information refers to the partial prosthesis used in a revision procedure. In general, partial arthroplasties are exchanged by a total knee prosthesis.

4.3. Characteristics of primary arthroplasties

An overall of 49,427 knee arthroplasties could be classified for the whole study period. Total knee arthroplasties represented 97.0%, whereas partial knee arthroplasties represented 3.0%. Within total knee arthroplasties, the most common arthroplasties were those which preserve the posterior cruciate retaining (CR) and the posterior stabilized ones (PS), representing 46.0% and 51.7%, respectively (Table 11). Among partial knee arthroplasties, unicompartmental arthroplasties were the most common group followed by patellofemoral arthroplasties, representing the 82.9% and 16.3% of partial procedures, respectively.

The most frequent primary knee arthroplasties across age groups were also CR and PS (Figure 8). Unicompartmental prosthesis was the most common type among partial procedures, especially among older patients.

The most common fixation was cemented fixation both for total and partial knee arthroplasties (Table 12, Table 13 and Figure 9). Again, in total and partial knee arthroplasties the most common fixation technique in men and women (Table 14, Table 15) and across age groups (Table 16 and Table 17) was cemented fixation.

In relation to the type of arthroplasty, the cemented fixation technique was the most used overall. CR and PS were cemented in 63.8% and 87.7% of cases, respectively (Table 18). For partial knee arthroplasties, cemented fixation was the most common fixation type both for unicompartmental and patellofemoral prosthesis (Table 19).

Figure 8. Percentage distribution of primary knee arthroplasties by age group. A) Total arthroplasties; B) Partial arthroplasties. Source RACat.




Table 11. Number of primary knee arthroplasties per type of arthroplasty and period.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|------------------------------------|--------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Total knee arthroplasties | | | | | | | | | | | | |
| Posterior cruciate retaining (CR) | 2,060 | 45.5 | 3,875 | 42.2 | 5,649 | 45.1 | 5,604 | 48.1 | 4,856 | 48.3 | 22,044 | 46,0 |
| Posterior stabilized (PS) | 2,403 | 53.0 | 5,061 | 55.1 | 6,579 | 52.5 | 5,842 | 50.1 | 4,892 | 48.7 | 24,777 | 51,7 |
| Constrained | 42 | 0.9 | 104 | 1.1 | 132 | 1.1 | 93 | 0.8 | 156 | 1.6 | 527 | 1,1 |
| Tumoral | 0 | 0.0 | 2 | 0.0 | 1 | 0.0 | 5 | 0.0 | 8 | 0.1 | 16 | 0,0 |
| Hinge | 3 | 0.1 | 24 | 0.3 | 44 | 0.4 | 25 | 0.2 | 40 | 0.4 | 136 | 0,3 |
| Rotational | 24 | 0.5 | 116 | 1.3 | 130 | 1.0 | 85 | 0.7 | 97 | 1.0 | 452 | 0,9 |
| Partial knee arthroplasties | | | | | | | | | | | | |
| Patellofemoral | 2 | 1.9 | 46 | 20.6 | 78 | 20.7 | 73 | 21.4 | 41 | 9.7 | 240 | 16,3 |
| Unicompartmental | 102 | 98.1 | 176 | 78.9 | 293 | 77.7 | 268 | 78.6 | 384 | 90.4 | 1223 | 82,9 |
| Bicompartmental | 0 | 0.0 | 1 | 0.5 | 6 | 1.6 | 0 | 0.0 | 0 | 0.0 | 7 | 0,5 |
| Other types | 1 | 100.0 | 4 | 100.0 | 13 | 100.0 | 8 | 100.0 | 12 | 100.0 | 33 | 100.0 |
| Not specified | 3,186 | 100.0 | 2,178 | 100.0 | 1,894 | 100.0 | 1,162 | 100.0 | 2,312 | 100.0 | 10,732 | 100.0 |
| Total | 7,823 | 100.0 | 11,587 | 100.0 | 14,819 | 100.0 | 13,165 | 100.0 | 12,798 | 100.0 | 60,192 | 100.0 |



Table 12. Number of primary total knee arthroplasties per fixation technique by time period. Source: RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|-----------------------|--------------|--------------|--------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Cemented | 3,235 | 71.4 | 6,841 | 74.5 | 9,314 | 74.3 | 8,999 | 77.2 | 7,901 | 78.6 | 36,290 | 75.7 |
| Cementless | 36 | 0.8 | 66 | 0.7 | 230 | 1.8 | 85 | 0.7 | 23 | 0.2 | 440 | 0.9 |
| Hybrid | 1,111 | 24.5 | 1,667 | 18.2 | 1,967 | 15.7 | 1,378 | 11.8 | 960 | 9.6 | 7,083 | 14.8 |
| Inverse hybrid | 0 | 0.0 | 50 | 0.5 | 131 | 1.1 | 99 | 0.9 | 72 | 0.7 | 352 | 0.7 |
| Not specified | 150 | 3.3 | 558 | 6.1 | 893 | 7.1 | 1,093 | 9.4 | 1,093 | 10.9 | 3,787 | 7.9 |
| Total | 4,532 | 100.0 | 9,182 | 100.0 | 12,535 | 100.0 | 11,654 | 100.0 | 10,049 | 100.0 | 47,952 | 100.0 |

Table 13. Number of primary partial knee arthroplasties per fixation technique by time period. Source: RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|----------------------|------------|--------------|------------|--------------|------------|--------------|------------|--------------|------------|--------------|--------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Cemented | 99 | 95.2 | 173 | 77.6 | 299 | 78.3 | 264 | 77.5 | 372 | 87.5 | 1,202 | 81.5 |
| Cementless | 3 | 2.9 | 3 | 1.4 | 3 | 0.8 | 2 | 0.6 | 0 | 0.0 | 11 | 0.8 |
| Not specified | 2 | 1.9 | 47 | 21.1 | 80 | 20.9 | 75 | 22.0 | 53 | 12.5 | 257 | 17.4 |
| Total | 104 | 100.0 | 223 | 100.0 | 382 | 100.0 | 341 | 100.0 | 425 | 100.0 | 1,475 | 100.0 |

Figure 9. Percentage distribution of the fixation technique in primary knee arthroplasties by time period. A) Total knee arthroplasties; B) Partial knee arthroplasties. Source: RACat



Prosthesis with not specified fixation technique were not plotted.

Table 14. Number of primary total knee arthroplasties per fixation technique by sex. Source RACat.

| | Men | | Women | | Total | |
|-----------------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % |
| Cemented | 9,998 | 72.8 | 26,292 | 76.8 | 36,290 | 75.7 |
| Cementless | 152 | 1.1 | 288 | 0.8 | 440 | 0.9 |
| Hybrid | 2,018 | 14.7 | 5,065 | 14.8 | 7,083 | 14.8 |
| Inverse hybrid | 68 | 0.5 | 284 | 0.8 | 352 | 0.7 |
| Not specified | 1,497 | 10.9 | 2,290 | 6.7 | 3,787 | 7.9 |
| Total | 13,733 | 100.0 | 34,219 | 100.0 | 47,952 | 100.0 |

Table 15. Number of primary partial knee arthroplasties per fixation technique by sex. Source RACat

| | Men | | Women | | Total | |
|----------------------|------------|--------------|------------|--------------|--------------|--------------|
| | n | % | n | % | n | % |
| Cemented | 412 | 84.6 | 795 | 80.5 | 1,207 | 81.8 |
| Cementless | 6 | 1.2 | 5 | 0.5 | 11 | 0.8 |
| Not specified | 69 | 14.2 | 188 | 19.0 | 257 | 17.4 |
| Total | 487 | 100.0 | 988 | 100.0 | 1,475 | 100.0 |

Table 16. Number of primary total knee arthroplasties per fixation technique by age group. Source RACat.

| | <65 | | 65-74 | | 75-84 | | >=85 | | Total | |
|-----------------------|--------------|--------------|---------------|--------------|---------------|--------------|------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % |
| Cemented | 5,848 | 71.6 | 15,371 | 74.7 | 14,298 | 78.3 | 773 | 80.8 | 36,290 | 75.7 |
| Cementless | 145 | 1.8 | 185 | 0.9 | 103 | 0.6 | 7 | 0.7 | 440 | 0.9 |
| Hybrid | 1,374 | 16.8 | 3,286 | 16.0 | 2,340 | 12.8 | 83 | 8.7 | 7,083 | 14.8 |
| Inverse hybrid | 107 | 1.3 | 170 | 0.8 | 75 | 0.4 | 0 | 0.0 | 352 | 0.7 |
| Not specified | 692 | 8.5 | 1,560 | 7.6 | 1,441 | 7.9 | 94 | 9.8 | 3,787 | 7.9 |
| Total | 8,166 | 100.0 | 20,572 | 100.0 | 18,257 | 100.0 | 957 | 100.0 | 47,952 | 100.0 |

Table 17. Number of primary partial knee arthroplasties per fixation technique by age group. Source RACat.

| | <65 | | 65-74 | | 75-84 | | >=85 | | Total | |
|----------------------|------------|--------------|------------|--------------|------------|--------------|-----------|--------------|--------------|--------------|
| | n | % | n | % | n | % | n | % | n | % |
| Cemented | 505 | 75.0 | 469 | 86.9 | 220 | 89.1 | 13 | 92.9 | 1,207 | 81.8 |
| Cementless | 7 | 1.0 | 1 | 0.2 | 3 | 1.2 | 0 | 0.0 | 11 | 0.8 |
| Not specified | 162 | 24.0 | 70 | 13.0 | 24 | 9.7 | 1 | 7.1 | 257 | 17.4 |
| Total | 674 | 100.0 | 540 | 100.0 | 247 | 100.0 | 14 | 100.0 | 1,475 | 100.0 |

Table 18. Number of primary total knee arthroplasties by fixation technique. Source RACat

| | CR | | PS | | Constrained | | Tumoral | | Hinge | | Rotational | | Total | |
|-----------------------|---------------|--------------|---------------|------------|-------------|--------------|-----------|--------------|------------|--------------|------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % | n | % |
| Cemented | 14,067 | 63.8 | 21,737 | 88 | 424 | 80.5 | 0 | 0.0 | 58 | 42.7 | 4 | 0.9 | 36,290 | 75.7 |
| Cementless | 292 | 1.3 | 148 | 1 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 440 | 0.9 |
| Inverse hybrid | 190 | 0.9 | 162 | 1 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 352 | 0.7 |
| Hybrid | 5,384 | 24.4 | 1,665 | 7 | 32 | 6.1 | 2 | 12.5 | 0 | 0.0 | 0 | 0.0 | 7,083 | 14.8 |
| Not specified | 2,111 | 9.6 | 1,065 | 4 | 71 | 13.5 | 14 | 87.5 | 78 | 57.4 | 448 | 99.1 | 3,787 | 7.9 |
| Total | 22,044 | 100.0 | 24,777 | 100 | 527 | 100.0 | 16 | 100.0 | 136 | 100.0 | 452 | 100.0 | 47,952 | 100.0 |

CR: Posterior cruciate retaining; PS: Posterior stabilized

Table 19. Number of primary partial knee arthroplasties by fixation technique. Source: RACat


| | Patellofemoral | | Bicompartmental | | Unicompartmental | | Total | |
|----------------------|----------------|--------------|-----------------|------------|------------------|--------------|--------------|--------------|
| | n | % | n | % | n | % | n | % |
| Cemented | 15 | 93.8 | 7 | 100 | 1,185 | 96.9 | 1,207 | 81.8 |
| Cementless | 0 | 0.0 | 0 | 0 | 11 | 0.9 | 11 | 0.8 |
| Not specified | 225 | 93.8 | 0 | 0 | 27 | 2.2 | 252 | 17.4 |
| Total | 240 | 100.0 | 7 | 100 | 1,223 | 100.0 | 1,470 | 100.0 |

4.4. Characteristics of implant models

In the period 2005-2014, from all the prosthesis correctly classified, a total of 137 models (counting the cemented and cementless models separately) were identified. Figure 10 shows the variability in the usage of different prosthesis models in the hospitals participating in the RACat. As a general trend, there is a tendency towards use of a wider range of models as the number of arthroplasties performed in the hospital rises. In primary knee arthroplasties, the number of models used per hospital ranged from 1 to 38, with an average of 12.

The five most frequent implant models by type of fixation and by type of knee arthroplasty are described in Table 20 and Table 21, respectively. NEXGEN was the most common cemented implant across the study period. Models for cementless and inverse hybrid prosthesis showed some variation over the study periods, GEMINI LINK being the most used model in both groups during the period 2013-2014. For hybrid prosthesis, ADVANCE model was used mainly during the period 2005-2012, but SIGMA increased during the latter period.

Figure 10. Number of different prosthesis models in relation to the volume of primary knee arthroplasties in the 2005-2014 period. Source RACat.



Both total and partial arthroplasties were considered.

Table 20. Five most frequent models in primary knee arthroplasty per fixation type and period. Source RACat.

| 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | |
|-----------------------|-----|------|-----------------|-------|------|---------------------|-------|------|-----------------------|-------|------|---------------------|-------|------|
| Model | n | % | Model | n | % | Model | n | % | Model | n | % | Model | n | % |
| Cemented | | | | | | | | | | | | | | |
| NEXGEN cim | 743 | 22.3 | NEXGEN cim | 1,735 | 24.7 | NEXGEN cim | 1,642 | 17.1 | NEXGEN cim | 1,680 | 18.1 | NEXGEN cim | 1,692 | 20.5 |
| PROFIX cim | 386 | 11.6 | SIGMA cim | 809 | 11.5 | PROFIX cim | 1,348 | 14.0 | PROFIX cim | 1,113 | 12.0 | GENUTECH cim | 1,170 | 14.1 |
| SCORPIO cim | 334 | 10.0 | PROFIX cim | 689 | 9.8 | GENUTECH cim | 1,015 | 10.6 | GENUTECH cim | 999 | 10.8 | SIGMA cim | 1,154 | 14.0 |
| ADVANCE cim | 312 | 9.4 | SCORPIO cim | 571 | 8.1 | SCORPIO NRG cim | 884 | 9.2 | SIGMA cim | 740 | 8.0 | TRIATHLON cim | 891 | 10.8 |
| OPTETRAK cim | 266 | 8.0 | GENUTECH cim | 416 | 5.9 | SIGMA cim | 833 | 8.7 | TRIATHLON cim | 921 | 10.0 | LEGION cim | 549 | 6.6 |
| Cementless | | | | | | | | | | | | | | |
| TRI CCC | 17 | 43.6 | TRI CCC | 22 | 31.9 | ALPINA | 97 | 41.6 | TRI CCC | 25 | 28.7 | GEMINI LINK | 11 | 47.8 |
| SKS | 11 | 28.2 | SKS | 21 | 30.4 | VANGUARD | 45 | 19.3 | VANGUARD ROCC | 16 | 18.4 | INNEX | 3 | 13.0 |
| EMOTION | 6 | 15.4 | EMOTION | 17 | 24.6 | TRI CCC | 30 | 12.9 | GEMINI LINK | 13 | 14.9 | PROFIX | 1 | 4.4 |
| UKS | 3 | 7.7 | NEXGEN | 4 | 5.8 | SKS | 22 | 9.4 | SKS | 10 | 11.5 | TRI CCC | 1 | 4.4 |
| PROFIX | 2 | 5.1 | UKS | 3 | 4.4 | GENESIS II | 13 | 5.6 | ALPINA | 8 | 9.2 | VANGUARD ROCC | 7 | 30.4 |
| Hybrid | | | | | | | | | | | | | | |
| ADVANCE | 696 | 62.6 | ADVANCE | 601 | 36.0 | ADVANCE | 688 | 34.9 | ADVANCE | 521 | 37.8 | SIGMA | 311 | 32.4 |
| 913 | 139 | 12.5 | AGC | 455 | 27.3 | AGC | 546 | 27.7 | GENUTECH | 332 | 24.1 | GENUTECH | 298 | 31.0 |
| GENUTECH | 89 | 8.0 | GENUTECH | 194 | 11.6 | SIGMA | 265 | 13.5 | SIGMA | 214 | 15.5 | ADVANCE | 172 | 17.9 |
| SIGMA | 77 | 6.9 | SIGMA | 106 | 6.4 | GENUTECH | 100 | 5.1 | AGC | 68 | 4.9 | UKNEE | 82 | 8.5 |
| UKNEE | 30 | 2.7 | 913 | 77 | 4.6 | VANGUARD | 72 | 3.7 | UKNEE | 61 | 4.4 | VANGUARD | 26 | 2.7 |
| Inverse hybrid | | | | | | | | | | | | | | |
| | | | GEMINI LINK cim | 35 | 70.0 | GEMINI LINK cim | 54 | 40.9 | NEXGEN cim | 38 | 36.9 | GEMINI LINK cim | 21 | 29.2 |
| | | | NEXGEN cim | 10 | 20.0 | NEXGEN cim | 50 | 37.9 | VANGUARD cim | 33 | 32.0 | NEXGEN LPS FLEX cim | 1 | 1.4 |
| | | | EMOTION cim | 1 | 2.0 | VANGUARD cim | 18 | 13.6 | GEMINI LINK cim | 25 | 24.3 | NEXGEN cim | 35 | 48.6 |
| | | | PROFIX cim | 1 | 2.0 | ALPINA cim | 3 | 2.3 | HLS UNI EVOLUTION cim | 3 | 2.9 | VANGUARD cim | 15 | 20.8 |
| | | | SCORPIO cim | 1 | 2.0 | NEXGEN LPS FLEX cim | 3 | 2.3 | PROFIX cim | 1 | 1.0 | | 0 | 0.0 |

Table 21. Five most frequent models in primary knee arthroplasty per type of arthroplasty and period. Source RACat.

| 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | |
|---------------------|-----|------|-----------------|-------|-------|-----------------|-------|-------|-----------------|-------|-------|---------------|-------|-------|
| Model | n | % | Model | n | % | Model | n | % | Model | n | % | Model | n | % |
| CR | | | | | | | | | | | | | | |
| ADVANCE | 400 | 19.4 | PROFIX cim | 710 | 18.3 | PROFIX cim | 1,313 | 23.2 | PROFIX cim | 1,090 | 19.5 | SIGMA cim | 977 | 20.1 |
| PROFIX cim | 385 | 18.7 | SIGMA cim | 619 | 16.0 | SIGMA cim | 610 | 10.8 | VANGUARD cim | 705 | 12.6 | VANGUARD cim | 692 | 14.3 |
| SIGMA cim | 170 | 8.3 | ADVANCE | 389 | 10.0 | ADVANCE | 425 | 7.5 | SIGMA cim | 558 | 10.0 | GENUTECH cim | 523 | 10.8 |
| 913 | 139 | 6.8 | AGC | 268 | 6.9 | GENUTECH cim | 338 | 6.0 | GENUTECH cim | 382 | 6.8 | TRIATHLON cim | 339 | 7.0 |
| SCORPIO cim | 129 | 6.3 | GENUTECH | 195 | 5.0 | AGC | 280 | 5.0 | GENUTECH | 326 | 5.8 | SIGMA | 317 | 6.5 |
| PS | | | | | | | | | | | | | | |
| NEXGEN cim | 665 | 27.7 | NEXGEN cim | 1,616 | 31.9 | NEXGEN cim | 1,556 | 23.7 | NEXGEN cim | 1,675 | 28.7 | NEXGEN cim | 1,688 | 34.5 |
| ADVANCE | 295 | 12.3 | SCORPIO cim | 485 | 9.6 | SCORPIO NRG cim | 942 | 14.3 | TRIATHLON cim | 662 | 11.3 | GENUTECH cim | 588 | 12.0 |
| SCORPIO cim | 208 | 8.7 | GENUTECH cim | 384 | 7.6 | GENUTECH cim | 715 | 10.9 | GENUTECH cim | 601 | 10.3 | TRIATHLON cim | 559 | 11.4 |
| OPTETRAK cim | 184 | 7.7 | SCORPIO NRG cim | 270 | 5.3 | TRIATHLON cim | 467 | 7.1 | GENESIS II cim | 393 | 6.7 | LEGION cim | 397 | 8.1 |
| ALPINA cim | 181 | 7.5 | SIGMA cim | 233 | 4.6 | GENESIS II cim | 333 | 5.1 | NEXGEN GSF cim | 357 | 6.1 | VANGUARD cim | 319 | 6.5 |
| Constrained | | | | | | | | | | | | | | |
| PERFORMANCE cim | 15 | 35.7 | NEXGEN cim | 45 | 43.3 | NEXGEN cim | 29 | 22.0 | NEXGEN cim | 27 | 29.0 | GENUTECH cim | 66 | 42.3 |
| NEXGEN cim | 8 | 19.1 | PERFORMANCE cim | 21 | 20.2 | GENUTECH cim | 27 | 20.5 | GENUTECH cim | 21 | 22.6 | TRIATHLON cim | 27 | 17.3 |
| SCORPIO cim | 7 | 16.7 | SCORPIO cim | 13 | 12.5 | SCORPIO cim | 20 | 15.2 | GENESIS II cim | 8 | 8.6 | NEXGEN cim | 23 | 14.7 |
| ADVANCE cim | 5 | 11.9 | 913 cim | 10 | 9.6 | PERFORMANCE cim | 16 | 12.1 | TRIATHLON cim | 7 | 7.5 | GENUTECH | 18 | 11.5 |
| SIGMA TC3 cim | 4 | 9.5 | GENUTECH cim | 7 | 6.7 | LEGION cim | 12 | 9.1 | PERFORMANCE cim | 6 | 6.5 | LEGION cim | 11 | 7.1 |
| Tumoral | | | PSO | 2 | 100.0 | STANMORE cim | 1 | 100.0 | METS | 5 | 100.0 | METS | 7 | 100.0 |
| Hinge | | | | | | | | | | | | | | |
| ENDO-MODEL LINK cim | 1 | 33.3 | ROTAX cim | 11 | 45.8 | ENDO-MODEL-M | 11 | 25.0 | ENDO-MODEL-M | 9 | 36.0 | ENDO-MODEL-M | 15 | 37.5 |

| 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | |
|-------------------------|----|-------|------------------------|-----|-------|------------------------|-----|-------|--------------------------|-----|------|----------------------------|-----|-------|
| Model | n | % | Model | n | % | Model | n | % | Model | n | % | Model | n | % |
| NOILES cim | 1 | 33.3 | ENDO-MODEL-M | 5 | 20.8 | NEXGEN cim | 8 | 18.2 | MRH cim | 3 | 12.0 | NEXGEN cim | 9 | 22.5 |
| ROTAX cim | 1 | 33.3 | GMRS cim | 2 | 8.3 | RHK cim | 8 | 18.2 | OSS cim | 3 | 12.0 | MEGASYS- TEM-C LINK cim | 8 | 20.0 |
| | | | NOILES cim | 2 | 8.3 | ROTAX cim | 7 | 15.9 | MEGASYSTEM-C LINK cim | 2 | 8.0 | OSS cim | 3 | 7.5 |
| | | | NEXGEN cim | 1 | 4.2 | MRH cim | 4 | 9.1 | METS cim | 2 | 8.0 | NOILES cim | 2 | 5.0 |
| Rotational | | | | | | | | | | | | | | |
| ENDO-MODEL LINK cim | 24 | 100.0 | ENDO-MODEL LINK cim | 116 | 100.0 | ENDO-MODEL LINK cim | 128 | 98.5 | ENDO-MODEL LINK cim | 83 | 97.7 | ENDO-MODEL LINK cim | 97 | 100.0 |
| | | | | | | RHK cim | 2 | 1.5 | PROFIX cim | 1 | 1.2 | | | |
| | | | | | | | | | RHK cim | 1 | 1.2 | | | |
| Patellofemoral | | | | | | | | | | | | | | |
| MIS AVON cim | 1 | 50.0 | ACCURIS cim | 2 | 4.4 | COMPETITOR cim | 36 | 46.2 | COMPETITOR cim | 47 | 64.4 | COMPETITOR cim | 24 | 58.5 |
| SPHEROCEN- TRIC cim | 1 | 50.0 | COMPETITOR cim | 34 | 73.9 | FPV cim | 13 | 16.7 | PFJ cim | 9 | 12.3 | VANGUARD PFC cim | 6 | 14.6 |
| Inverse hybrid | | | MIS AVON cim | 5 | 10.9 | ACCURIS cim | 8 | 10.3 | FPV cim | 6 | 8.2 | FPV cim | 5 | 12.2 |
| | | | SPHEROCENTRIC cim | 4 | 8.7 | PFJ cim | 5 | 6.4 | ACCURIS cim | 5 | 6.9 | PFJ cim | 2 | 4.9 |
| | | | VANGUARD PFC cim | 1 | 2.2 | | | | SPHEROCENTRIC cim | 2 | 2.7 | | | |
| Bicompartmental | | | | | | | | | | | | | | |
| | | | JOURNEY DEUCE cim | 1 | 100.0 | JOURNEY DEUCE cim | 6 | 100.0 | | | | | | |
| Unicompartmental | | | | | | | | | | | | | | |
| ZIMMER UNI KNEE cim | 26 | 25.5 | ZIMMER UNI KNEE cim | 54 | 30.7 | ACCURIS cim | 110 | 37.5 | ACCURIS cim | 156 | 58.2 | ACCURIS cim | 253 | 65.9 |
| ACCURIS cim | 25 | 24.5 | ACCURIS cim | 50 | 28.4 | OXFORD III cim | 76 | 25.9 | OXFORD III cim | 54 | 20.2 | SIGMA cim | 64 | 16.7 |

| 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | |
|------------------------|----|------|---------------------|----|------|---------------------|----|------|---------------------|----|-----|----------------|----|------|
| Model | n | % | Model | n | % | Model | n | % | Model | n | % | Model | n | % |
| OXFORD III cim | 16 | 15.7 | OXFORD III cim | 42 | 23.9 | ZIMMER UNI KNEE cim | 54 | 18.4 | TRIATHLON cim | 21 | 7.8 | OXFORD III cim | 51 | 13.3 |
| MILLER/GAL-LAN- TE cim | 13 | 12.8 | ENDO-MODEL LINK cim | 14 | 8.0 | ENDO-MODEL LINK cim | 15 | 5.1 | SIGMA cim | 11 | 4.1 | TRIATHLON cim | 6 | 1.6 |
| ADVANCE cim | 10 | 9.8 | EIUS cim | 7 | 4.0 | COMPETITOR cim | 13 | 4.4 | ZIMMER UNI KNEE cim | 9 | 3.4 | OXFORD cim | 4 | 1.0 |

Percentages were calculated for each model out of total prosthesis used in each type of arthroplasty. Only the values of the 5 most common models are shown in the table, which explains that for some types of arthroplasties percentages do not add up to 100.



CR: Posterior cruciate retaining; PS: Posterior stabilized

4.5. Implant survival in knee arthroplasties

Follow-up description

The median length of follow-up from the date of primary arthroplasty until the revision surgery or censorship was 4.22 years (SD: 2.48) and 3.48 years (SD: 2.37) for total and partial arthroplasties, respectively. Figure 11 shows the distribution of follow-up time for patients undergoing total and partial knee arthroplasty.

Figure 11. Distribution of the follow-up time of the patients with primary knee arthroplasties. A) Total; B) Partial. Source: RACat.



Cumulative revision rate

The estimated cumulative revision rate for total and partial knee arthroplasties (with no adjustment) is shown in Table 22 and Table 23. For total knee arthroplasties, the 1-year revision rate was 0.89%, rising to 3.8% at 5-year and 5.1% at 9-years after the primary surgery. For partial knee arthroplasties, the rate was 2.0% and 10.2% at the 1-year and 5-year after the primary surgery, respectively.

The risk of revision both for patients who underwent total and partial knee arthroplasty was adjusted by sex, age, and comorbidities. A higher risk of revision was observed in groups of younger patients. The Cox model fitted for total knee arthroplasties revealed an increased risk of revision for patients under 65 years and for patients aged between 65 and 75 years compared to those aged 85 years or over (HR=3.34 (IC95%: 1.96-5.70); HR=2.02 (IC95%: 1.19-3.43); data not shown). A similar trend was observed for partial knee arthroplasties, although differences were not statistically significant (data not shown).

The cumulative revision rate by fixation technique (without adjustment) for total knee arthroplasties revealed that cemented and hybrid prostheses showed the lowest revision rate at 1-year (0.8% and 0.7%, respectively), whereas the 1-year revision rate for cementless prosthesis was 3.0% (Table 24, Figure 12). For partial knee arthroplasties, the cumulative revision rate for cemented prosthesis was 2.0% at 1-year follow up, rising to 7.4% and 10.1% at 3- and 5-year follow-up, respectively (Table 25). Estimating the risk of revision of total knee arthroplasties per fixation technique using a Cox model (adjusted by sex, age and comorbidities), revealed a higher risk of revision of cementless and inverse hybrid arthroplasties in comparison with cemented arthroplasties (HR=2.00 (IC95%: 1.43-2.81); HR=1.58 (IC95%: 1.01-2.46), respectively). However, the risk of revision was lower among prosthesis using hybrid fixation (HR=0.69 (IC95%: 0.59-0.81; data not shown).

Table 26 shows the cumulative revision rates for CR and PS, the two most common types of total knee arthroplasty. The data shows that the cumulative rate at 1- and 3-year of follow-up is quite similar between both arthroplasty types. However, the 9-year cumulative rate was somewhat higher for PS in comparison to CR. The cumulative revision rate for partial knee arthroplasties revealed that 1-year revision rate for unicompartmental arthroplasties was 2.1%, increasing up to 7.0% and 9.3% at 3- and 5-year follow-up, respectively (Table 27). Risk of revision adjusted by sex, age and comorbidities for CR and PS arthroplasties revealed no differences in the risk of revision for both types of arthroplasties (data not shown).

We estimated the cumulative incidence of replacement considering death as a competitive event, both for partial and total knee arthroplasties. Figure 13 shows two functions: one with the cumulative risk of revision adjusted by age, sex, comorbidity and patient death probability, and the other adjusted by the same variables without considering the competitive risk of death. It can be observed that both curves are quite similar, the cumulative incidence function without adjustment by competitive risk of death slowly growing apart from the other curve. Even though the difference is not very significant, it increases over time and 4-5 years after the intervention, the difference becomes more evident. For total knee arthroplasty, the cumulative incidence of revision, adjusted by age, sex, and comorbidity at 5 years was 4.02% and considering competitive risk of death, 3.99%. For partial knee arthroplasties, the cumulative incidence of revision, adjusted by age, sex, and comorbidity at 5 years was 10.74% and considering competitive risk of death, 10.65%

Table 22. Estimated cumulative revision rate after a primary total knee arthroplasty. Source: RACat-MBD-SHD-RCA

| Time | Primary arthroplasty | Revisions | Cumulative revision rate (%) | CI 95% | |
|---------|----------------------|-----------|------------------------------|--------|------|
| 30 days | 45,019 | 76 | 0.17 | 0.13 | 0.21 |
| 90 days | 43,934 | 101 | 0.39 | 0.34 | 0.46 |
| 1 year | 40,289 | 209 | 0.89 | 0.81 | 0.98 |
| 3 years | 28,832 | 671 | 2.77 | 2.61 | 2.94 |
| 5 years | 16,812 | 266 | 3.84 | 3.63 | 4.05 |
| 7 years | 7,546 | 91 | 4.53 | 4.29 | 4.79 |
| 9 years | 790 | 26 | 5.17 | 4.76 | 5.63 |

95%CI: 95% confidence interval. RACat: Catalan Arthroplasty Register

Table 23. Estimated cumulative revision rate after a primary partial knee arthroplasty. Source: RACat-MBD-SHD-RCA.

| Time | Primary arthroplasty | Revisions | Cumulative Revision rate (%) | CI 95% | |
|---------|----------------------|-----------|------------------------------|--------|-------|
| 30 days | 1,387 | 1 | 0.07 | 0.01 | 0.51 |
| 90 days | 1,339 | 1 | 0.15 | 0.04 | 0.58 |
| 1 year | 1,134 | 22 | 2.07 | 1.41 | 3.02 |
| 3 years | 706 | 54 | 7.55 | 6.1 | 9.34 |
| 5 years | 391 | 18 | 10.24 | 8.42 | 12.43 |
| 7 years | 136 | 4 | 11.6 | 9.44 | 14.22 |
| 9 years | 12 | 2 | . | . | . |

95%CI: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

Table 24. Estimated cumulative revision rate after a primary total knee arthroplasty by fixation technique.
Source: RACat- MBDSHD-RCA

| Time | Primary arthroplasty | Revisions | Cumulative Revision rate (%) | CI 95% | |
|-----------------------|----------------------|-----------|------------------------------|--------|-------|
| Cemented | | | | | |
| 30 days | 34,145 | 60 | 0.18 | 0.14 | 0.23 |
| 90 days | 33,296 | 77 | 0.40 | 0.34 | 0.47 |
| 1 year | 30,428 | 143 | 0.86 | 0.77 | 0.97 |
| 3 years | 21,380 | 515 | 2.8 | 2.61 | 3 |
| 5 years | 12,113 | 207 | 3.92 | 3.69 | 4.18 |
| 7 years | 5,535 | 73 | 4.7 | 4.4 | 5.01 |
| 9 years | 379 | 22 | 5.62 | 4.95 | 6.37 |
| Cementless | | | | | |
| 30 days | 400 | 0 | 0.00 | . | . |
| 90 days | 400 | 0 | 0.00 | . | . |
| 1 year | 385 | 12 | 3.02 | 1.73 | 5.26 |
| 3 years | 316 | 17 | 7.5 | 5.27 | 10.62 |
| 5 years | <i>171</i> | 6 | 9.5 | 6.89 | 13.02 |
| 7 years | 51 | 0 | . | . | . |
| 9 years | 1 | 0 | . | . | . |
| Hybrid | | | | | |
| 30 days | 6,669 | 6 | 0.09 | 0.04 | 0.2 |
| 90 days | 6,564 | 8 | 0.21 | 0.12 | 0.35 |
| 1 year | 6,176 | 30 | 0.68 | 0.51 | 0.91 |
| 3 years | 5,014 | 87 | 2.18 | 1.84 | 2.58 |
| 5 years | 3,417 | 32 | 2.87 | 2.46 | 3.34 |
| 7 years | 1,622 | 12 | 3.31 | 2.84 | 3.85 |
| 9 years | 385 | 4 | 3.61 | 3.07 | 4.25 |
| Inverse Hybrid | | | | | |
| 30 days | 344 | 0 | 0.00 | . | . |
| 90 days | 336 | 1 | 0.29 | 0.04 | 0.21 |
| 1 year | 313 | 5 | 1.83 | 0.83 | 4.04 |
| 3 years | 211 | 10 | 5.68 | 3.5 | 9.16 |
| 5 years | <i>109</i> | 4 | 7.93 | 5.1 | 12.23 |
| 7 years | 12 | 0 | . | . | . |
| 9 years | 1 | 0 | . | . | . |

95%CI: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

Figure 12. Estimated cumulative revision rate after a primary total knee arthroplasty by fixation technique.
 Source: RACat-MBDSHD-RCA




Table 25. Estimated cumulative revision rate after a primary partial knee arthroplasty by fixation technique.
 Source: RACat- MBDSHD-RCA

| Time | Primary arthroplasty | Revisions | Cumulative Revision rate (%) | CI 95% | |
|-----------------|----------------------|-----------|------------------------------|--------|-------|
| Cemented | | | | | |
| 30 days | 1,348 | 1 | 0.07 | 0.01 | 0.52 |
| 90 days | 1,301 | 1 | 0.15 | 0.04 | 0.6 |
| 1 year | 1,101 | 23 | 2.05 | 1.39 | 3.02 |
| 3 years | 687 | 52 | 7.45 | 5.99 | 9.25 |
| 5 years | 384 | 17 | 10.05 | 8.23 | 12.25 |
| 7 years | 132 | 4 | 11.45 | 9.27 | 14.1 |
| 9 years | 12 | 2 | . | . | . |

95%CI: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

The number of patients at risk for the other fixation categories was <100; the estimates were not calculated since they were very unreliable.

Table 26. Estimated cumulative revision rate after a primary total knee arthroplasty (CR/PS)

| Time, in years | Primary arthroplasty | Revisions | Cumulative Revision rate (%) | CI 95% | |
|----------------|----------------------|-----------|------------------------------|--------|------|
| CR | | | | | |
| 30 days | 21,017 | 25 | 0.12 | 0.08 | 0.18 |
| 90 days | 20,477 | 47 | 0.34 | 0.27 | 0.43 |
| 1 year | 18,633 | 40 | 0.91 | 0.78 | 1.05 |
| 3 years | 13,155 | 306 | 2.76 | 2.53 | 3.02 |
| 5 years | 7,516 | 113 | 3.78 | 3.48 | 4.1 |
| 7 years | 3,444 | 36 | 4.40 | 4.04 | 4.78 |
| 9 years | 382 | 12 | 4.79 | 4.38 | 5.24 |
| PS | | | | | |
| 30 days | 23,168 | 45 | 0.19 | 0.14 | 0.26 |
| 90 days | 22,660 | 49 | 0.41 | 0.33 | 0.5 |
| 1 year | 20,918 | 94 | 0.84 | 0.73 | 0.96 |
| 3 years | 15,140 | 350 | 2.73 | 2.51 | 2.97 |
| 5 years | 8,969 | 145 | 3.81 | 3.54 | 4.11 |
| 7 years | 3,996 | 54 | 4.59 | 4.25 | 4.96 |
| 9 years | 405 | 14 | 5.46 | 4.77 | 6.24 |

95%CI: 95% confidence interval. RACat: Catalan Arthroplasty Register

CR: Posterior cruciate retaining; PS: Posterior stabilized



Table 27. Estimated cumulative revision rate after a primary partial knee arthroplasty (Unicompartmental/Patellofemoral)

| Time | Primary arthroplasty | Revisions | Cumulative Revision rate (%) | CI 95% | |
|-------------------------|----------------------|-----------|------------------------------|-------------|--------------|
| Unicompartmental | | | | | |
| 30 days | 1,146 | 1 | 0.09 | 0.01 | 0.62 |
| 90 days | 1,102 | 1 | 0.18 | 0.04 | 0.7 |
| 1 year | 914 | 18 | 2.13 | 1.41 | 3.22 |
| 3 years | 551 | 38 | 6.99 | 5.45 | 8.95 |
| 5 years | 328 | 12 | 9.27 | 7.36 | 11.64 |
| 7 years | 121 | 4 | <i>10.88</i> | <i>8.54</i> | <i>13.82</i> |
| 9 years | 12 | 1 | . | . | . |
| Patellofemoral | | | | | |
| 30 days | 229 | 0 | 0.00 | . | . |
| 90 days | 226 | 0 | 0.00 | . | . |
| 1 year | 211 | 2 | 0.92 | 0.23 | 3.64 |
| 3 years | 146 | 16 | 9.34 | 5.98 | 14.44 |
| 5 years | 57 | 5 | . | . | . |
| 7 years | 17 | 0 | . | . | . |
| 9 years | 1 | 1 | . | . | . |

CI 95%: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

Figure 13. Cumulative incidence of revision after a primary total and partial knee arthroplasty adjusted by competitive risk of death, sex and age. A) Total knee arthroplasty; B) Partial knee arthroplasty. Source: RACat-RCA



5. HIP ARTHROPLASTIES RESULTS

5.1. Scope and quality of the data

In the period described, data from 46,488 patients with primary hip arthroplasty and from 5,237 patients with revision hip arthroplasty was received. The volume of data increased over time, suffering a slight decrease from 2012 onwards (Figure 14). The highest volume of primary and revision arthroplasties was registered in the health care region of Barcelona (Table 28).

Table 29 shows the volume of primary and revision arthroplasties sent during the study period, per centre. The overall burden of revision, understood as the percentage of revision arthroplasties compared to the total number of events sent, was 11.3%.

Figure 15 shows the completeness both for primaries and revisions over the study period. The data shows that the highest completeness was reached in 2011 and 2012. A gradual decrease was observed from 2012 onwards, coinciding with technical changes in data notification. Figure 16 shows the variability in data completeness of each centre (blue points) compared with overall completeness (green line) and a confidence interval based on overall completeness. The centres that are below the range of the confidence interval exhibit lower completeness compared to the overall figure, and the ones that are above exhibit higher completeness than the overall. Twelve out of the 53 centres fell below the confidence interval based on the overall reference value.

As shown in Table 30, in the latter period 77.1% of hip arthroplasties were classified. Figure 17 is a funnel plot which shows the percentage of classified primary arthroplasties per centre in relation to the volume of data sent in the period 2013-2014. It can be observed great variability between centres on the percentage of classified cases, most of them presenting percentages around 80%, which indicates good quality of the information sent.

The percentage of arthroplasties with reported laterality increased over the study period reaching in 2013-2014 a 97.6% (Table 31). Using data from the latest 2 years we plotted percentage of laterality reported per centre (Figure 18). In that case, the percentage of completeness for this variable is high in most centres.

Figure 14. Volume of hip arthroplasties sent to the RACat per year. Source: RACat




Table 28. Number of hip arthroplasties by health region, period and by type of arthroplasty. Source RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | All periods | |
|---------------------|--------------|------------|--------------|------------|---------------|-------------|---------------|-------------|---------------|-------------|---------------|--------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| Lleida | 90 | 19 | 248 | 26 | 332 | 33 | 535 | 28 | 587 | 16 | 1,792 | 122 |
| Camp de Tarragona | 86 | 1 | 458 | 26 | 771 | 49 | 745 | 57 | 709 | 71 | 2,769 | 204 |
| Terres de l'Ebre | 0 | 1 | 0 | 0 | 275 | 23 | 276 | 18 | 324 | 24 | 875 | 66 |
| Girona | 924 | 80 | 1,217 | 133 | 1,429 | 168 | 1,435 | 159 | 1,423 | 107 | 6,428 | 647 |
| Catalunya Central | 438 | 58 | 880 | 94 | 972 | 134 | 1,065 | 101 | 1,141 | 97 | 4,496 | 484 |
| Alt Pirineu i Aragó | 116 | 5 | 104 | 7 | 111 | 8 | 122 | 6 | 103 | 4 | 556 | 30 |
| Barcelona | 3,520 | 398 | 5,118 | 604 | 6,951 | 931 | 7,232 | 956 | 6,751 | 795 | 29,572 | 3,684 |
| Total | 5,174 | 562 | 8,025 | 890 | 10,841 | 1346 | 11,410 | 1325 | 11,038 | 1114 | 46,488 | 5,237 |

Table 29. Number of hip arthroplasties by health care region, centre, period and by type of arthroplasty. Source RACat.

| Hospital | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|---|-----------|----------|-----------|----------|-----------|----------|-----------|----------|-----------|----------|--------------|------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| Lleida | | | | | | | | | | | | |
| Hospital Universitari Arnau de Vilanova | 0 | 0 | 0 | 0 | 78 | 14 | 303 | 7 | 309 | 0 | 690 | 21 |
| Hospital de Santa Maria | 70 | 16 | 171 | 13 | 153 | 13 | 108 | 14 | 178 | 7 | 680 | 63 |
| Clínica de Ponent | 20 | 3 | 77 | 13 | 101 | 6 | 124 | 7 | 100 | 9 | 422 | 38 |
| Camp de Tarragona | | | | | | | | | | | | |
| Hospital Universitari de Tarragona Joan XXIII | 0 | 0 | 147 | 12 | 211 | 24 | 207 | 33 | 214 | 30 | 779 | 99 |
| Hospital Universitari Sant Joan de Reus | 0 | 0 | 70 | 0 | 302 | 0 | 293 | 0 | 122 | 0 | 787 | 0 |
| Hospital de Sant Pau i Santa Tecla | 19 | 0 | 64 | 3 | 79 | 16 | 81 | 11 | 123 | 8 | 366 | 38 |
| Pius Hospital de Valls | 48 | 1 | 106 | 3 | 100 | 6 | 92 | 6 | 94 | 4 | 440 | 20 |
| Hospital del Vendrell | 19 | 0 | 71 | 8 | 79 | 3 | 72 | 7 | 156 | 29 | 397 | 47 |
| Terres de l'Ebre | | | | | | | | | | | | |
| Hospital Tortosa Verge de la Cinta | 0 | 1 | 0 | 0 | 275 | 23 | 219 | 18 | 270 | 24 | 764 | 66 |
| Hospital Comarcal Móra d'Ebre | 0 | 0 | 0 | 0 | 0 | 0 | 57 | 0 | 54 | 0 | 111 | 0 |
| Girona | | | | | | | | | | | | |
| Hospital Universitari de Girona Josep Trueta | 0 | 0 | 235 | 49 | 270 | 52 | 261 | 46 | 284 | 9 | 1,050 | 156 |
| Clínica Girona SA | 0 | 0 | 0 | 0 | 39 | 4 | 73 | 7 | 0 | 0 | 112 | 11 |
| Clínica Salus Infirmorum | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 42 | 1 | 42 | 1 |
| Hospital Comarcal de Blanes | 298 | 24 | 306 | 36 | 298 | 24 | 270 | 15 | 325 | 24 | 1,497 | 123 |
| Hospital de Campdevanòl | 38 | 2 | 41 | 1 | 41 | 0 | 48 | 1 | 52 | 0 | 220 | 4 |
| Hospital de Figueres | 107 | 8 | 257 | 15 | 278 | 32 | 247 | 38 | 113 | 11 | 1,002 | 104 |
| Hospital de Palamós | 204 | 20 | 226 | 16 | 193 | 21 | 195 | 23 | 212 | 34 | 1,030 | 114 |
| Hospital de Sant Jaume d'Olot | 146 | 16 | 36 | 2 | 131 | 13 | 150 | 18 | 164 | 14 | 627 | 63 |
| Hospital Provincial Santa Caterina | 131 | 10 | 116 | 14 | 179 | 22 | 191 | 11 | 231 | 14 | 848 | 71 |
| Catalunya Central | | | | | | | | | | | | |
| Centre Hospitalari-ALTHAIA | 203 | 42 | 452 | 63 | 451 | 98 | 470 | 80 | 506 | 71 | 2,082 | 354 |
| Hospital General de Vic | 129 | 2 | 161 | 1 | 237 | 3 | 314 | 2 | 400 | 0 | 1,241 | 8 |
| Hospital de Sant Bernabé | 0 | 0 | 41 | 3 | 74 | 9 | 85 | 2 | 107 | 15 | 307 | 29 |
| Fundació Sanitària d'Igualada FP | 106 | 14 | 226 | 27 | 210 | 24 | 196 | 17 | 128 | 11 | 866 | 93 |
| Alt Pirineu i Aran | | | | | | | | | | | | |
| Fundació Sant Hospital de la Seu d'Urgell | 32 | 2 | 35 | 0 | 44 | 1 | 45 | 3 | 39 | 2 | 195 | 8 |
| Hospital Comarcal del Pallars | 30 | 2 | 34 | 6 | 26 | 3 | 27 | 2 | 8 | 2 | 125 | 15 |
| Hospital de Puigcerdà | 54 | 1 | 35 | 1 | 41 | 4 | 50 | 1 | 56 | 0 | 236 | 7 |

| Hospital | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|--|--------------|------------|--------------|------------|---------------|--------------|--------------|--------------|---------------|--------------|---------------|--------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| Barcelona | | | | | | | | | | | | |
| Hospital de Viladecans | 0 | 0 | 0 | 0 | 74 | 21 | 160 | 21 | 190 | 4 | 424 | 46 |
| Hospital Universitari de Bellvitge | 358 | 0 | 287 | 0 | 438 | 8 | 203 | 15 | 194 | 5 | 1,480 | 28 |
| H.Universitari Germans Trias i Pujol | 0 | 0 | 10 | 3 | 186 | 37 | 237 | 18 | 334 | 3 | 767 | 61 |
| Hospital Clinic i Provincial de Barcelona | 703 | 109 | 639 | 68 | 780 | 133 | 701 | 126 | 593 | 132 | 3,416 | 568 |
| Hospital de L'Hospitalet - CSI | 275 | 29 | 227 | 23 | 337 | 43 | 303 | 53 | 366 | 22 | 1,508 | 170 |
| Hospital Dos de Maig de Barcelona - CSI | 75 | 6 | 64 | 14 | 150 | 18 | 187 | 21 | 201 | 15 | 677 | 74 |
| Hospital Mútua de Terrassa | 0 | 0 | 256 | 51 | 363 | 74 | 304 | 54 | 90 | 21 | 1,013 | 200 |
| Hospital de l'Esperit Sant | 61 | 31 | 133 | 21 | 232 | 21 | 199 | 23 | 198 | 39 | 823 | 135 |
| Corporació Sanitària Parc Taulí | 184 | 35 | 431 | 96 | 451 | 92 | 422 | 92 | 276 | 61 | 1,764 | 376 |
| Hospital de Sant Boi- Parc Sanitari St Joan de Deu | 92 | 7 | 134 | 5 | 145 | 13 | 170 | 20 | 223 | 19 | 764 | 64 |
| Hospital de Sant Celoni - Fundació privada | 70 | 5 | 118 | 3 | 93 | 5 | 81 | 1 | 138 | 2 | 500 | 16 |
| Hospital de Terrassa | 0 | 0 | 184 | 36 | 223 | 39 | 211 | 23 | 259 | 40 | 877 | 138 |
| Parc de Salut del Mar | 95 | 12 | 0 | 0 | 506 | 57 | 674 | 110 | 0 | 0 | 1,275 | 179 |
| Hospital General de Granollers | 191 | 25 | 237 | 25 | 66 | 5 | 335 | 23 | 237 | 66 | 1,066 | 144 |
| Hospital Municipal de Badalona | 160 | 14 | 167 | 10 | 132 | 14 | 160 | 14 | 112 | 11 | 731 | 63 |
| Hospital Residència Sant Camil | 192 | 35 | 188 | 51 | 207 | 30 | 212 | 27 | 204 | 20 | 1,003 | 163 |
| Hospital Sant Joan de Déu de Martorell | 70 | 14 | 161 | 51 | 195 | 22 | 161 | 21 | 139 | 12 | 726 | 120 |
| Hospital de Sant Joan de D'Esplugues Llob | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 | 0 | 0 | 0 | 1 |
| Hospital Sant Rafael | 0 | 0 | 0 | 0 | 179 | 53 | 155 | 30 | 134 | 30 | 468 | 113 |
| F.G.S. Hospital de la Santa Creu i Sant Pau | 0 | 0 | 1 | 0 | 362 | 44 | 664 | 82 | 728 | 90 | 1,755 | 216 |
| H. de Plató | 0 | 0 | 87 | 6 | 98 | 11 | 35 | 3 | 68 | 4 | 288 | 24 |
| Hospital Universitari Sagrat Cor | 244 | 32 | 392 | 43 | 371 | 40 | 380 | 45 | 402 | 102 | 1,789 | 262 |
| Fundació Privada Hospital de Mollet | 129 | 6 | 95 | 3 | 102 | 7 | 175 | 7 | 199 | 3 | 700 | 26 |
| Hospital Comarcal de l'Alt Penedes | 1 | 0 | 146 | 30 | 163 | 20 | 184 | 24 | 177 | 26 | 671 | 100 |
| Hospital de Mataro | 145 | 38 | 281 | 64 | 366 | 56 | 327 | 62 | 373 | 38 | 1,492 | 258 |
| H. Universitari Vall d'Hebron | 475 | 0 | 880 | 1 | 732 | 68 | 592 | 40 | 916 | 30 | 3,595 | 139 |
| Total | 5,174 | 562 | 8,025 | 890 | 10,841 | 1,346 | 1,141 | 1,325 | 11,038 | 1,114 | 46,488 | 5,237 |

Figure 15. Completeness of data sent to the RACat per year, and by type of arthroplasty. Source RACat-MBD-SHD




Figure 16. Percentage of completeness of primary hip arthroplasties per centre during the period 2013-2014. Source: RACat




Table 30. Number and percentages of classified primary hip arthroplasties reported by period. Source RACat

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|-----------------------|--------------|--------------|--------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Classified | 2,916 | 56.4 | 5,872 | 73.2 | 8,889 | 82.0 | 9,839 | 86.2 | 8,509 | 77.1 | 36,025 | 77.5 |
| Not classified | 2,258 | 43.6 | 2,153 | 26.8 | 1,952 | 18.0 | 1,571 | 13.8 | 2,529 | 22.9 | 10,463 | 22.5 |
| Total | 5,174 | 100.0 | 8,025 | 100.0 | 10,841 | 100.0 | 11,410 | 100.0 | 11,038 | 100.0 | 46,488 | 100.0 |

Figure 17. Percentage of primary hip arthroplasties classified per centre according to the volume of data sent in the period 2013-2014. Source: RACat.




Table 31. Number of hip arthroplasties with the operated side (left or right) reported by period. Source RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|-------------------------------|--------------|--------------|--------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Reported operated side | 2,378 | 41.5 | 1,806 | 20.3 | 541 | 4.4 | 73 | 0.6 | 296 | 2.4 | 5094 | 9.9 |
| Not reported | 3,358 | 58.5 | 7,109 | 79.7 | 11,646 | 95.6 | 12,662 | 99.4 | 11,856 | 97.6 | 46,631 | 90.2 |
| Total | 5,736 | 100.0 | 8,915 | 100.0 | 12,187 | 100.0 | 12,735 | 100.0 | 12,152 | 100.0 | 51,725 | 100.0 |

Figure 18. Percentage of primary and revision hip arthroplasty with operated side informed per centre, and in relation to data volume sent in the period 2013-2014. Source: RACat




Table 32. Age and sex of patients undergoing a primary total hip arthroplasty by health care region and period. Source: RACat

| Health care region | 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | | Total | | |
|--------------------|--------------|--------------------|-------------|--------------|--------------------|-------------|--------------|---------------------|-------------|--------------|--------------------|-------------|--------------|--------------------|-------------|---------------|--------------------|-------------|
| | n | Mean age (SD) | % women | n | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | n | Mean age (SD) | % women |
| Lleida | 80 | 69.2 (10.8) | 40.0 | 183 | 67.1 (11.8) | 42.1 | 259 | 66.8 (11.7) | 46.7 | 254 | 70.2 (11.0) | 46.5 | 269 | 69.5 (11.6) | 44.2 | 1,045 | 68.9 (11.6) | 44.9 |
| Camp de Tarragona | 45 | 70.9 (9.2) | 51.1 | 241 | 67.2 (12.2) | 50.6 | 339 | 67.7 (13.0) | 45.7 | 317 | 68.0 (11.7) | 46.1 | 398 | 67.4 (12.0) | 49.8 | 1,340 | 67.8 (12.2) | 48.0 |
| Terres de l'Ebre | 0 | 0 | 0.0 | 0 | 0 | 0.0 | 181 | 67.1 (13.5) | 47.0 | 108 | 65.9 (13.4) | 50.9 | 126 | 66.2 (10.8) | 50.8 | 415 | 68.5 (12.7) | 49.0 |
| Girona | 493 | 70.7 (10.4) | 56.2 | 591 | 70.6 (11.2) | 56.0 | 656 | 70.3 (12.0) | 55.3 | 625 | 69.8 (12.0) | 51.0 | 704 | 69.3 (11.2) | 50.6 | 3,069 | 70.1 (11.5) | 53.6 |
| Catalunya centra | 164 | 70.1 (10.7) | 48.2 | 451 | 69.4 (10.7) | 49.5 | 502 | 68.6 (12.8) | 50.0 | 567 | 68.1 (11.9) | 47.3 | 628 | 68.1 (11.5) | 46.5 | 2,312 | 68.7 (11.8) | 48.4 |
| Alt Pirineu i Aran | 64 | 73.7 (9.4) | 60.9 | 72 | 71.7 (9.9) | 55.6 | 73 | 75.2 (11.0) | 63.0 | 58 | 70.7 (11.4) | 43.1 | 65 | 70.5 (10.7) | 49.2 | 332 | 72.5 (10.5) | 54.7 |
| Barcelona | 1,175 | 68.2 (11.5) | 54.8 | 2,448 | 68.5 (12.3) | 54.2 | 3,733 | 68.0 (12.6) | 53.2 | 4,199 | 68.5 (12.4) | 53.6 | 3,244 | 67.8 (12.6) | 51.5 | 14,799 | 68.3 (12.5) | 53.3 |
| Total | 2,021 | 69.3 (11.1) | 54.1 | 3,986 | 68.8 (12.0) | 53.2 | 5,743 | 68.3 (112.6) | 52.3 | 6,128 | 68.6 (12.3) | 51.9 | 5,434 | 68.1 (12.2) | 50.3 | 23,312 | 68.5 (11.2) | 52.1 |

Table 33. Age and sex of patients undergoing a primary partial hip arthroplasty by health care region and period. Source: RACat

| Health care region | 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | | Total | | |
|--------------------|------------|-------------------|-------------|--------------|-------------------|-------------|--------------|-------------------|-------------|--------------|-------------------|-------------|--------------|-------------------|-------------|---------------|-------------------|-------------|
| | n | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | N | Mean age (SD) | % women | n | Mean age (SD) | % women |
| Lleida | 4 | 86.6 (4.9) | 100.0 | 7 | 86.1 (9.9) | 100.0 | 12 | 84.6 (3.6) | 83.3 | 170 | 84.6 (7.8) | 81.8 | 128 | 85.1 (6.7) | 74.2 | 321 | 84.8 (7.2) | 79.4 |
| Camp de Tarragona | 36 | 84.6 (7.5) | 80.6 | 98 | 84.0 (8.3) | 73.5 | 100 | 83.8 (8.2) | 69.0 | 114 | 85.1 (6.8) | 76.3 | 143 | 85.6 (7.5) | 70.6 | 491 | 84.7 (7.6) | 72.9 |
| Terres de l'Ebre | 0 | 0 | 0.0 | 0 | 0 | 0.0 | 86 | 84.5 (5.1) | 83.7 | 103 | 83.8 (6.5) | 78.6 | 91 | 85.3 (5.6) | 79.1 | 280 | 84.5 (5.8) | 80.4 |
| Girona | 140 | 85.8 (6.2) | 78.6 | 90 | 86.5 (5.2) | 81.1 | 212 | 84.9 (7.1) | 75.9 | 320 | 84.8 (6.6) | 79.1 | 356 | 84.9 (7.8) | 75.6 | 1,118 | 85.1 (7.0) | 77.5 |
| Catalunya central | 114 | 83.4 (6.5) | 76.3 | 294 | 83.8 (6.7) | 75.5 | 353 | 84.0 (7.4) | 77.6 | 429 | 84.6 (7.4) | 73.0 | 411 | 84.8 (6.9) | 74.0 | 1,601 | 84.3 (7.1) | 75.0 |
| Alt Pirineu i Aran | 16 | 83.2 (8.4) | 68.8 | 23 | 86.8 (6.2) | 87.0 | 31 | 86.7 (7.0) | 67.7 | 59 | 84.5 (9.1) | 76.3 | 32 | 84.0 (7.4) | 81.3 | 161 | 85.5 (7.9) | 76.4 |
| Barcelona | 585 | 84.0 (7.4) | 76.9 | 1,374 | 84.2 (7.2) | 78.2 | 2,352 | 84.4 (7.2) | 74.9 | 2,516 | 85.3 (7.0) | 75.8 | 1,914 | 85.4 (7.3) | 74.7 | 8,741 | 84.6 (7.2) | 75.8 |
| Total | 895 | 84.2 (7.1) | 77.2 | 1,886 | 84.3 (7.1) | 77.8 | 3,146 | 84.4 (7.2) | 75.3 | 3,711 | 85.1 (7.1) | 76.1 | 3,075 | 85.3 (7.3) | 74.7 | 12,713 | 84.8 (7.2) | 75.9 |

Figure 19. Age of patients undergoing a hip arthroplasty by period. A) Total hip arthroplasty; B) Partial hip arthroplasty. Source: RACat





Table 34. Demographic and patient treatment profile of patients with primary and revision total hip arthroplasty by period. Source RACat-MBDSHD

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | |
|---------------------------------|-----------------|----------------|-----------------|----------------|-----------------|-----------------|-----------------|----------------|-----------------|----------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) |
| Number | 1,844 (100) | 78 (100) | 3,765 (100) | 118 (100) | 5,453 (100) | 246 (100) | 5,958 (100) | 207 (100) | 5,211 (100) | 128 (100) |
| % of Women | 1,002 (4.3) | 47 (60.3) | 1,996 (53.0) | 72 (61.0) | 2,857 (52.4) | 1,133 (54.1) | 3,085 (51.8) | 115 (55.6) | 2,615 (50.2) | 69 (53.9) |
| Mean age (SD) | 69.3 (11.0) | 73.0 (12.2) | 68.8 (12.0) | 73.7 (12.1) | 68.4 (12.5) | 73.4 (11.2) | 68.5 (12.2) | 72.7 (12.4) | 68.0 (12.1) | 72.9 (12.2) |
| Cause of intervention | | | | | | | | | | |
| Osteoarthritis | 1,431 (77.6) | 0 | 3,008 (79.9) | 0 | 4,418 (81.0) | 0 | 4,810 (80.7) | 0 | 4,256 (81.7) | 0 |
| Fracture | 228 (12.4) | 0 | 394 (10.5) | 0 | 565 (10.4) | 0 | 654 (11.0) | 0 | 481 (9.2) | 0 |
| Other diagnostics | 185 (10.0) | 0 | 363 (9.6) | 0 | 470 (8.6) | 0 | 494 (8.3) | 0 | 474 (9.1) | 0 |
| Cause of revision | | | | | | | | | | |
| Mechanical complications | 0 | 54 (69.2) | 0 | 81 (68.6) | 0 | 170 (69.1) | 0 | 149 (72.0) | 0 | 80 (62.5) |
| Infection | 0 | 9 (11.5) | 0 | 8 (6.8) | 0 | 30 (12.2) | 0 | 21 (10.1) | 0 | 21 (16.4) |
| Other diagnostics | 0 | 15 (19.2) | 0 | 29 (24.6) | 0 | 46 (18.7) | 0 | 37 (17.9) | 0 | 27 (21.1) |
| Comorbidity | | | | | | | | | | |
| 1 or more | 955 (51.8) | 44 (56.4) | 1,927 (51.2) | 66 (55.9) | 3,205 (58.8) | 166 (67.5) | 3,752 (63.0) | 139 (67.2) | 3,329 (63.9) | 88 (68.8) |
| Healthcare discharge | 162 (8.8) | 9 (11.5) | 317 (8.4) | 21 (17.8) | 562 (10.3) | 54 (22.0) | 763 (12.8) | 44 (21.3) | 652 (12.5) | 31 (24.2) |

Table 35. Demographic and patient treatment profile of patients with primary and revision partial hip arthroplasty by period. Source RACat-MBDSHD

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | |
|---------------------------------|---------------|---------------|-----------------|---------------|-----------------|---------------|-----------------|---------------|-----------------|---------------|
| | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision | Primary | Revision |
| | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) | n (%) |
| Number | 790 (100) | 2 (100) | 1,817 (100) | 13 (100) | 2,953 (100) | 18 (100) | 3,659 (100) | 16 (100) | 2,973 (100) | 13 (100) |
| % of Women | 612 (77.5) | 2 (100) | 1,419 (78.1) | 7 (53.9) | 2,222 (75.3) | 14 (77.8) | 2,788 (76.2) | 10 (62.5) | 2,228 (74.9) | 11 (84.6) |
| Mean age (SD) | 84.2 (7.0) | 91.3 (3.1) | 84.4 (7.0) | 79.8 (8.0) | 84.5 (7.2) | 82.8 (9.4) | 85.1 (7.1) | 85.0 (7.1) | 85.3 (7.2) | 84.7 (6.0) |
| Cause of intervention | | | | | | | | | | |
| Fracture | 752 (95.2) | 0 | 1,724 (94.9) | 0 | 2,799 (94.8) | 0 | 3,504 (95.8) | 0 | 2,841 (95.6) | 0 |
| Other diagnostics | 38 (4.8) | 0 | 93 (5.1) | 0 | 154 (5.2) | 0 | 152 (4.3) | 0 | 132 (4.4) | 0 |
| Cause of revision | | | | | | | | | | |
| Mechanical complications | 0 | 1 (50.0) | 0 | 7 (53.9) | 0 | 14 (77.8) | 0 | 10 (62.5) | 0 | 6 (46.2) |
| Infection | 0 | 0 | 0 | 4 (30.8) | 0 | 1 (5.6) | 0 | 2 (12.5) | 0 | 6 (46.2) |
| Other diagnostics | 0 | 1 (50.0) | 0 | 2 (15.4) | 0 | 3 (16.7) | 0 | 4 (25.0) | 0 | 1 (7.69) |
| Comorbidity | | | | | | | | | | |
| 1 or more | 560 (70.9) | 1 (50.0) | 1,356 (74.6) | 7 (53.9) | 2,289 (77.5) | 16 (88.9) | 3,047 (83.3) | 14 (87.5) | 2,057 (84.3) | 11 (84.6) |
| Healthcare discharge | 191 (24.2) | 1 (50.0) | 475 (26.1) | 0 (0.0) | 945 (32.0) | 3 (16.7) | 1,342 (36.7) | 6 (37.5) | 1,253 (42.2) | 6 (46.6) |

5.3. Characteristics of primary arthroplasties

30,031 hip arthroplasties could be classified for the whole study period. Total hip arthroplasties represented 77.6% whereas partial arthroplasties represented 22.4%. Within total hip arthroplasties, the most common implants were the conventional prosthesis, representing over 95% of arthroplasties. Bipolar and unipolar monoblock were the most common types of partial prosthesis, representing 46.7% and 47.2% of partial arthroplasties, respectively (Table 36).

Conventional total hip arthroplasties were the most common arthroplasty type across age groups. However, for partial hip arthroplasties, the use of bipolar prosthesis decreased proportionally as age increased, whereas the opposite pattern was observed for unipolar monoblock prosthesis (Figure 20).

Almost 65% of total hip arthroplasties were cementless, this type of fixation showing an increasing trend over time (Figure 21). The proportion of cemented and cementless prosthesis for partial hip arthroplasties was 49.5% and 50.5%, respectively. (Table 37, Table 38, and Figure 21). Among total hip arthroplasties, the percentage of cementless prosthesis was higher for men than for women (73.1% and 57.4%, respectively) (Table 39). Regarding partial hip arthroplasties, the distribution of type of fixation for men and women was quite similar (Table 40).

Table 41 and Table 42 show the distribution of type of fixation technique by age groups, for total and partial arthroplasties, respectively. For total fixation techniques, the data shows that the percentage of cementless prostheses decreased proportionally as age increased. However, for partial hip arthroplasties, the older the patients, the higher the percentage of cementless prostheses.

In conventional total hip arthroplasties, the most commonly used friction device was metal-polyethylene (65.8%). This type of friction device was the most common in all age groups, even though there was more variability in the friction device used among patients under 65 (Table 43).

Head size distribution for total conventional hip arthroplasties by friction devices revealed that small heads (<32mm) were commonly used among Metal on Polyethylene and Ceramic on Polyethylene implants (72.7% and 68.1%, respectively). Medium heads were the most common among Ceramic on Ceramic arthroplasties (71%). Among Metal on Metal prosthesis, small heads represented 43.3% and large heads 35.7% (Table 44).

Table 36. Number of primary hip arthroplasties per type of arthroplasty and period. Source: RACat.

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|---------------------------------|--------------|--------------|--------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Total Hip Arthroplasty | | | | | | | | | | | | |
| Double mobility | 0 | 0.00 | 0 | 0.00 | 2 | 0.03 | 14 | 0.2 | 15 | 0.3 | 31 | 0.1 |
| Resurfacing | 23 | 1.1 | 124 | 3.1 | 234 | 4.1 | 121 | 2.0 | 46 | 0.9 | 548 | 2.4 |
| Conventional | 1,996 | 98.8 | 3,835 | 96.2 | 5,382 | 93.7 | 5,841 | 95.3 | 5,176 | 95.3 | 22,230 | 95.4 |
| Short stem | 2 | 0.1 | 27 | 0.7 | 125 | 2.2 | 152 | 2.5 | 197 | 3.6 | 503 | 2.2 |
| Partial Hip Arthroplasty | | | | | | | | | | | | |
| Bipolar | 339 | 37.9 | 818 | 43.4 | 1,617 | 51.4 | 1,668 | 45 | 1,491 | 48.5 | 5,933 | 46.7 |
| Unipolar modular | 96 | 10.7 | 132 | 7.0 | 231 | 7.3 | 207 | 5.6 | 114 | 3.7 | 780 | 6.14 |
| Unipolar monoblock | 460 | 51.4 | 936 | 49.6 | 1,298 | 41.3 | 1,836 | 49.5 | 1,470 | 47.8 | 6,000 | 47.2 |
| Not specified | 2,258 | 100.0 | 2,153 | 100.0 | 1,952 | 100.0 | 1,571 | 100.0 | 2,529 | 100.0 | 10,463 | 100.0 |
| Total | 5,174 | 100.0 | 8,025 | 100.0 | 10,841 | 100.0 | 11,410 | 100.0 | 11,038 | 100.0 | 46,488 | 100.0 |

Figure 20. Percentage distribution of primary hip arthroplasties by age group and type of arthroplasty. A) Total hip arthroplasty; B) Partial hip arthroplasty. Source RACat.




Table 37. Number of primary total hip arthroplasties per fixation technique by time period. Source: RACat

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|-----------------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Cemented | 324 | 16.0 | 438 | 11.0 | 497 | 8.7 | 654 | 10.7 | 483 | 8.9 | 2,396 | 10.3 |
| Cementless | 1,065 | 52.7 | 2,408 | 60.4 | 3,798 | 66.1 | 4,049 | 66.1 | 3,809 | 70.1 | 15,129 | 64.9 |
| Hybrid | 571 | 28.3 | 990 | 24.8 | 1,214 | 21.1 | 1,243 | 20.3 | 997 | 18.4 | 5,015 | 21.5 |
| Inverse hybrid | 61 | 3.0 | 150 | 3.8 | 230 | 4.0 | 176 | 2.9 | 130 | 2.4 | 747 | 3.2 |
| Not specified | 0 | 0.0 | 0 | 0.0 | 4 | 0.1 | 6 | 0.1 | 15 | 0.3 | 25 | 0.1 |
| Total | 2,021 | 100.0 | 3,986 | 100.0 | 5,743 | 100.0 | 6,128 | 100.0 | 5,434 | 100.0 | 23,312 | 100.0 |

Table 38. Number of primary partial hip arthroplasties per fixation technique by time period. Source: RACat

| | 2005/2006 | | 2007/2008 | | 2009/2010 | | 2011/2012 | | 2013/2014 | | Total | |
|----------------------|------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % | n | % |
| Cemented | 479 | 53.5 | 972 | 51.5 | 1,694 | 53.9 | 1,728 | 46.6 | 1,420 | 46.2 | 6,293 | 49.5 |
| Cementless | 416 | 46.5 | 914 | 48.5 | 1,452 | 46.2 | 1,983 | 53.4 | 1,654 | 53.8 | 6,419 | 50.5 |
| Not specified | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 1 | 0.0 | 1 | 0.0 |
| Total | 895 | 100.0 | 1,886 | 100.0 | 3,146 | 100.0 | 3,711 | 100.0 | 3,075 | 100.0 | 12,713 | 100.0 |

Table 39. Number of primary total hip arthroplasties per fixation technique and sex. Source RACat.

| | Men | | Women | | Total | |
|-----------------------|---------------|--------------|---------------|--------------|---------------|--------------|
| | n | % | n | % | n | % |
| Cemented | 677 | 6.1 | 1,719 | 14.2 | 2,396 | 10.3 |
| Cementless | 8,165 | 73.1 | 6,964 | 57.4 | 15,129 | 64.9 |
| Hybrid | 1,805 | 16.2 | 3,210 | 26.5 | 5,015 | 21.5 |
| Inverse hybrid | 510 | 4.6 | 237 | 2.0 | 747 | 3.2 |
| Not specified | 20 | 0.2 | 5 | 0.0 | 25 | 0.1 |
| Total | 11,177 | 100.0 | 12,135 | 100.0 | 23,312 | 100.0 |

Table 40. Number of primary partial hip arthroplasties per fixation technique and sex. Source RACat.

| | Men | | Women | | Total | |
|----------------------|--------------|--------------|--------------|--------------|---------------|--------------|
| | n | % | n | % | n | % |
| Cemented | 1,542 | 50.3 | 4,751 | 49.2 | 6,293 | 49.5 |
| Cementless | 1,521 | 49.6 | 4,898 | 50.8 | 6,419 | 50.5 |
| Not specified | 1 | 0.0 | 0 | 0.0 | 1 | 0.0 |
| Total | 3,064 | 100.0 | 9,649 | 100.0 | 12,713 | 100.0 |

Table 41. Number of primary total hip arthroplasties per fixation technique by age group. Source RACat.

| | <65 years | | 65-74 years | | 75-84 years | | ≥85 years | | Total | |
|-----------------------|--------------|--------------|--------------|--------------|--------------|--------------|------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % |
| Cemented | 99 | 1.3 | 363 | 5.0 | 1,611 | 21.7 | 323 | 35.2 | 2,396 | 10.3 |
| Cementless | 6,682 | 86.0 | 5,021 | 69.7 | 3,179 | 42.8 | 247 | 26.9 | 15,129 | 64.9 |
| Hybrid | 468 | 6.0 | 1,725 | 23.9 | 2,509 | 33.8 | 313 | 34.1 | 5,015 | 21.5 |
| Inverse hybrid | 500 | 6.4 | 92 | 1.3 | 121 | 1.6 | 34 | 3.7 | 747 | 3.2 |
| Not specified | 17 | 0.2 | 4 | 0.1 | 4 | 0.1 | 0 | 0.0 | 25 | 0.1 |
| Total | 7,766 | 100.0 | 7,205 | 100.0 | 7,424 | 100.0 | 917 | 100.0 | 23,312 | 100.0 |

Figure 21. Percentage distribution of the fixation technique of primary hip arthroplasties by time period.
 A) Total arthroplasties; B) Partial arthroplasties. Source RACat





Table 42. Number of primary partial hip arthroplasties per fixation technique by age group. Source RACat

| | <65 years | | 65-74 years | | 75-84 years | | ≥85 years | | Total | |
|----------------------|------------|--------------|-------------|--------------|--------------|--------------|--------------|--------------|---------------|--------------|
| | n | % | n | % | n | % | n | % | n | % |
| Cemented | 148 | 77.1 | 560 | 74.2 | 3,041 | 59.9 | 2,544 | 38.1 | 6,293 | 49.5 |
| Cementless | 43 | 22.4 | 195 | 25.8 | 2,039 | 40.1 | 4,142 | 62.0 | 6,419 | 50.5 |
| Not specified | 1 | 0.5 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 1 | 0.0 |
| Total | 192 | 100.0 | 755 | 100.0 | 5,080 | 100.0 | 6,686 | 100.0 | 12,713 | 100.0 |

Table 43. Number of total conventional primary hip arthroplasties by type of friction device per age groups. Source RACat

| | <65 years | | 65-74 years | | 75-84 years | | ≥85 years | | Total | |
|--------------------------------|--------------|------------|--------------|------------|--------------|------------|------------|------------|---------------|-------------|
| | n | % | n | % | n | % | n | % | n | % |
| Metal on Metal | 544 | 8.0 | 238 | 3.4 | 241 | 3.3 | 56 | 6.2 | 1,079 | 4.9 |
| Metal on Polyethylene | 2,867 | 42.0 | 5,000 | 70.4 | 6,026 | 81.5 | 741 | 81.3 | 14,634 | 65.8 |
| Ceramic on Ceramic | 1,301 | 19.1 | 328 | 4.6 | 104 | 1.4 | 6 | 0.7 | 1,739 | 7.8 |
| Ceramic on Polyethylene | 1,321 | 19.4 | 861 | 12.1 | 413 | 5.6 | 24 | 2.6 | 2,619 | 11.8 |
| Ceramic on Metal | 102 | 1.5 | 35 | 0.5 | 7 | 0.1 | 2 | 0.2 | 146 | 0.7 |
| Metal on Polycarbonate | 1 | 0.01 | 3 | 0.04 | 2 | 0.03 | 0 | 0 | 6 | 0.03 |
| Not specified | 691 | 10.1 | 637 | 9.0 | 597 | 8.1 | 82 | 9.0 | 2,007 | 9.0 |
| Total | 6,827 | 100 | 7,102 | 100 | 7,390 | 100 | 911 | 100 | 22,230 | 100 |

Table 44. Distribution of friction device per head size among conventional hip arthroplasties. Source RACat

| | Metal on Metal | | Metal on Polyethylene | | Ceramic on Ceramic | | Ceramic on Polyethylene | | Ceramic on Metal | |
|------------------|----------------|------------|-----------------------|------------|--------------------|------------|-------------------------|------------|------------------|-------------|
| | n | % | n | % | n | % | n | % | n | % |
| <32 mm | 467 | 43.3 | 10,639 | 72.7 | 409 | 23.5 | 1,773 | 68.1 | 67 | 45.9 |
| 32-36mm | 227 | 21.0 | 3,892 | 26.6 | 1,234 | 71.0 | 830 | 31.9 | 73 | 50.0 |
| >36 mm | 385 | 35.7 | 99 | 0.7 | 96 | 5.5 | 1 | 0.04 | 6 | 4.1 |
| Total | 1,079 | 100 | 14,630 | 100 | 1,739 | 100 | 2,604 | 100 | 146 | 100 |


Information on the head size was not available in 2,032 cases, representing 9.1% of conventional hip arthroplasties.

5.4. Characteristics of implant models

In the period 2005-2014, a total of 1,114 models of stem-socket combinations (counting separately the cemented and uncemented models. Ex. Evoris cim and Evoris) were identified. In primary total hip arthroplasties, the number of acetabular models used by the hospitals ranged from 1 to 34, with an average of 16 per hospital. As for stem models, the number ranged from 1 to 33 with an average of 15 models used per hospital. As observed for knee arthroplasties, the number of implanted models grows in proportion to an increase in the volume of arthroplasties per centre (Figure 22).

Table 45 shows the five most frequent models in primary hip arthroplasty per type of arthroplasty and period. In the case of total hip arthroplasties, each model is presented as the combination of stem and socket component, and in the partial hip arthroplasties the stem model and the manufacturer are specified.

Figure 22. Number of different prostheses models in relation to the volume of primary hip arthroplasties for each centre sent to the RACat in the 2013-2014 period. Source: RACat



Both total and partial arthroplasties were considered.

Table 45. Five most frequent models in primary hip arthroplasty per type of arthroplasty and period. Source: RACat

| 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | |
|-----------------------------------|-----|------|----------------------------|-----|------|-------------------------------|-----|-------|-------------------------------|-----|------|-----------------------------|-----|------|
| Model | n | % | Model | n | % | Model | n | % | Model | n | % | Model | n | % |
| Double mobility (stem/cup) | | | | | | | | | | | | | | |
| | | | | | | FURLONG/ NOVAE E | 2 | 100.0 | FURLONG/ POLARCUP | 3 | 21.4 | SHINE-C cim/ DMX | 8 | 53.3 |
| | | | | | | | | | A-ACUTA S/ POLARCUP | 2 | 14.3 | FURLONG/ DOUBLE MOBILITY | 2 | 13.3 |
| | | | | | | | | | AUTOBLOQUEANTE cim/NOVAE E | 1 | 7.1 | POLARSTEM/ POLARCUP | 2 | 13.3 |
| | | | | | | | | | EXETER cim/ NOVAE E | 1 | 7.1 | SHINE-C cim/ DUAL cim | 2 | 13.3 |
| | | | | | | | | | FURLONG/ NOVAE E | 1 | 7.1 | FURLONG/ DOUBLE MOBILITY | 1 | 6.7 |
| Resurfacing (head/cup) | | | | | | | | | | | | | | |
| ASR/ASR | 13 | 56.5 | ASR/ASR | 37 | 29.8 | RECAP/RECAP | 63 | 26.9 | RECAP/RECAP | 33 | 27.3 | CONSERVE PLUS/ CONSERVE | 23 | 50.0 |
| CONSERVE PLUS/ CONSERVE | 5 | 21.7 | RECAP/RECAP | 33 | 26.6 | CONSERVE PLUS/ CONSERVE | 47 | 20.1 | DUROM/ MMC | 25 | 20.7 | ICON TM/ICON TM | 13 | 28.3 |
| BHR/BHR | 4 | 17.4 | CONSERVE PLUS/ CONSERVE | 18 | 14.5 | CORMET/ CORMET | 45 | 19.2 | CONSERVE PLUS/ CONSERVE | 23 | 19.0 | RECAP/RECAP | 7 | 15.2 |
| RECAP/RECAP | 1 | 4.4 | ICON TM/ICON TM | 16 | 12.9 | ADEPT/ADEPT | 22 | 9.4 | CORMET/CORMET | 18 | 14.9 | CORMET/CORMET | 2 | 4.4 |
| | | | ADEPT/ ADEPT | 8 | 6.5 | BHR/BHR | 15 | 6.4 | ICON TM/ICON TM | 13 | 10.7 | BHR/BHR | 1 | 2.2 |
| Conventional (stem/cup) | | | | | | | | | | | | | | |
| VERSYS/TRILOGY | 210 | 10.5 | VERSYS cim/ TRILOGY | 358 | 9.3 | FURLONG/CSF PLUS | 305 | 5.7 | FURLONG/CSF PLUS | 450 | 7.7 | FURLONG/CSF PLUS | 410 | 7.9 |
| VERSYS cim/ TRILOGY | 158 | 7.9 | VERSYS/ TRILOGY | 356 | 9.3 | VERSYS cim/ TRILOGY | 277 | 5.2 | UNITED U2/ U2 | 339 | 5.8 | CORAIL/PINNACLE | 276 | 5.3 |

| 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | |
|--------------------------------------|-----|------|--------------------------------------|-----|------|--|-----|------|--|-----|------|-------------------------------------|-----|------|
| Model | n | % | Model | n | % | Model | n | % | Model | n | % | Model | n | % |
| FURLONG/FURLONG | 136 | 6.8 | UNITED U2/ UNITED U2 | 198 | 5.2 | UNITED U2/ UNITED U2 | 270 | 5.0 | VERSYS cim/ TRILOGY | 246 | 4.2 | VERSYS cim/TRILOGY | 213 | 4.1 |
| UNITED U2/ UNITED U2 | 134 | 6.7 | SYMAX/ TRIDENT | 152 | 4.0 | VERSYS/TRILOGY | 244 | 4.5 | CORAIL/PINNACLE | 235 | 4.0 | NOVATION ELEMENT/ NOVATION CROWN | 199 | 3.8 |
| ECO cim/EXPANSYS | 92 | 4.6 | CORAIL/PINNACLE | 110 | 2.9 | CORAIL/ PINNACLE | 236 | 4.4 | EXETER cim/ CONTEMPORARY cim | 195 | 3.3 | UNITED U2/UNITED U2 | 164 | 3.2 |
| Short stem (stem/cup) | | | | | | | | | | | | | | |
| MAYO/TRILOGY | 1 | 50.0 | PROXIMA/ASR | 18 | 66.7 | PROXIMA/ PINNACLE | 47 | 37.6 | PROXIMA/PINNACLE | 51 | 33.6 | ACCOLADE II/TRIDENT | 45 | 22.8 |
| PROXIMA/ASR | 1 | 50.0 | PROXIMA/ PINNACLE | 7 | 25.9 | C.F.P. LINK/ T.O.P. LINK | 27 | 21.6 | TAPERLOC MICROPLASTY/ EXCEED ABT | 30 | 19.7 | TAPERLOC MICROPLASTY/ BIHAPRO | 23 | 11.7 |
| | | | PROXIMA/ REFLECTION SP3 | 1 | 3.7 | C.F.P. LINK/ CFP-BETA CUP | 25 | 20.0 | MINIHIP/ TRABECULAR METAL | 24 | 15.8 | C.F.P. LINK/COMBICUP | 21 | 10.7 |
| | | | PROXIMA/TRIDENT | 1 | 3.7 | PROXIMA/ASR | 11 | 8.8 | C.F.P. LINK/ T.O.P. LINK | 15 | 9.9 | FURLONG EVOLUTION/CSF PLUS | 19 | 9.6 |
| | | | | | | TAPERLOC MICROPLASTY/ EXCEED ABT | 3 | 2.4 | C.F.P. LINK/ COMBICUP | 8 | 5.3 | PROFEMUR PRESERVE/ PROCOTYL | 13 | 6.6 |
| Bipolar (stem/shell) | | | | | | | | | | | | | | |
| AUTOBLOQUEANTE cim/BIARTICULAR | 57 | 16.8 | AUTOBLOQUEANTE cim/BIARTICULAR | 268 | 32.8 | AUTOBLOQUEANTE cim/BIARTICULAR | 354 | 21.9 | AUTOBLOQUEANTE cim/BIARTICULAR | 349 | 20.9 | AUTOBLOQUEANTE cim/BIARTICULAR | 324 | 21.7 |
| MULLER cim/UHR BIPOLAR | 46 | 13.6 | MAINSTREAM MULLER cim/ CAPTIVE | 120 | 14.7 | VERSYS cim/ MULTIPOLAR | 138 | 8.5 | KAREY cim/ MOONSTOONE | 157 | 9.4 | KAREY cim/ MOONSTOONE | 177 | 11.9 |
| MAINSTREAM MULLER cim/ CAPTIVE | 45 | 13.3 | VERSYS cim/ MULTIPOLAR | 89 | 10.9 | EXETER cim/ UHR BIPOLAR | 107 | 6.6 | VERSYS cim/ MULTIPOLAR | 156 | 9.4 | VERSYS cim/ MULTIPOLAR | 129 | 8.7 |
| ECO cim/CAPTIVE | 35 | 10.3 | EXETER cim/ UHR BIPOLAR | 63 | 7.7 | KAREY cim/ BIARTICULAR | 85 | 5.3 | SHINE-C cim/ BIARTICULAR | 117 | 7.0 | SHINE-C cim/ BIARTICULAR | 124 | 8.3 |
| VERSYS cim/ MULTIPOLAR | 35 | 10.3 | ECO cim/CAPTIVE | 60 | 7.3 | PMB cim/BIPOLAR | 85 | 5.3 | EXETER cim/UHR BIPOLAR | 77 | 4.6 | EXETER cim/ UHR BIPOLAR | 107 | 7.2 |

| 2005/2006 | | | 2007/2008 | | | 2009/2010 | | | 2011/2012 | | | 2013/2014 | | |
|--|-----|------|-----------------------------------|-----|------|-----------------------------------|-----|------|-----------------------------------|-------|------|-----------------------------------|-----|------|
| Model | n | % | Model | n | % | Model | n | % | Model | n | % | Model | n | % |
| Unipolar modular (stem/head) | | | | | | | | | | | | | | |
| ECO cim/HEMIHEAD | 60 | 62.5 | ECO cim/HEMIHEAD | 88 | 66.7 | AUTOBLOQUEANTE cim/ENDOCEFÁLICA | 75 | 32.5 | AUTOBLOQUEANTE cim/ENDOCEFÁLICA | 81 | 39.1 | TWINSYS cim/HEMIHEAD | 34 | 29.8 |
| ECO cim/S30 | 36 | 37.5 | FULLFIX cim/HEMIHEAD | 43 | 32.6 | CCM cim/HEMIHEAD | 56 | 24.2 | TWINSYS cim/HEMIHEAD | 54 | 26.1 | TAPERLOC cim/BIOMET Head | 30 | 26.3 |
| | | | MAINSTREAM MULLER cim/BIOMET Head | 1 | 0.8 | FULLFIX cim/HEMIHEAD | 45 | 19.5 | CCM cim/HEMIHEAD | 44 | 21.3 | AUTOBLOQUEANTE cim/ENDOCEFÁLICA | 22 | 19.3 |
| | | | | | | CEMIRA cim/HEMIHEAD | 27 | 11.7 | CEMIRA cim/HEMIHEAD | 25 | 12.1 | TAPERLOC/BIOMET Head | 18 | 15.8 |
| | | | | | | ECO cim/BIOMET Head | 21 | 9.1 | WAGNER SL/HEMIHEAD | 2 | 1.0 | MULLER cim/ENDOCEFÁLICA | 8 | 7.0 |
| Unipolar monoblock (stem/manufacture) | | | | | | | | | | | | | | |
| AUSTIN MOORE/ SURGIVAL C.O., S.A. | 203 | 44.1 | AUSTIN MOORE/ SURGIVAL C.O., S.A. | 392 | 41.9 | AUSTIN MOORE/ SURGIVAL C.O., S.A. | 595 | 45.8 | AUSTIN MOORE/ SURGIVAL C.O., S.A. | 1,000 | 54.5 | AUSTIN MOORE/ SURGIVAL C.O., S.A. | 899 | 61.2 |
| AUSTIN MOORE/ STRIKER IBERIA, SL | 111 | 24.1 | AUSTIN MOORE/ STRYKER IBERIA, SL | 357 | 38.1 | AUSTIN MOORE/ STRYKER IBERIA, SL. | 518 | 39.9 | AUSTIN MOORE/ STRYKER IBERIA, SL. | 648 | 35.3 | AUSTIN MOORE/ STRYKER IBERIA, SL. | 450 | 30.6 |
| THOMPSON cim/ SURGIVAL C.O., S.A. | 76 | 16.5 | THOMPSON cim/ SURGIVAL C.O., S.A. | 84 | 9.0 | THOMPSON/ STRYKER IBERIA, SL. | 89 | 6.9 | THOMPSON/ STRYKER IBERIA, SL. | 98 | 5.3 | THOMPSON/ STRYKER IBERIA, SL. | 79 | 5.4 |
| AUSTIN MOORE/ BIOMET | 46 | 10.0 | AUSTIN MOORE/ TRAIBER, S.L. | 36 | 3.9 | THOMPSON cim/ SURGIVAL C.O., S.A. | 80 | 6.2 | THOMPSON cim/ SURGIVAL C.O., S.A. | 78 | 4.3 | THOMPSON cim/ SURGIVAL C.O., S.A. | 29 | 2.0 |
| AUSTIN MOORE/ TRAIBER, S.L. | 24 | 5.2 | AUSTIN MOORE/ BIOMET | 35 | 3.7 | THOMPSON cim/ STRYKER IBERIA, SL. | 7 | 0.5 | THOMPSON cim/ STRYKER IBERIA, SL. | 6 | 0.3 | THOMPSON cim/ BIOMET | 11 | 0.8 |


Percentages were calculated for each type of prosthesis out of total of prostheses used in each type of arthroplasty. Only the values of the 5 most common models are shown in the table, which explains that for some types of arthroplasties percentages do not add up to 100.

5.5. Implant survival in hip arthroplasties

Follow-up description

The median length of follow-up from the date of primary arthroplasty until the revision surgery or censorship was 3.93 years (SD: 2.42) and 3.71 years (SD: 2.28) for total and partial hip arthroplasties, respectively. Figure 23 shows the distribution of follow-up time for patients who underwent total and partial hip arthroplasty.

Figure 23. Distribution of the follow-up time of patients undergoing a hip arthroplasty. A) Total arthroplasty; B) Partial arthroplasty. Source: RACat



Cumulative revision rate

The estimated cumulative revision rate after a primary total and partial hip arthroplasty (with no adjustment) is shown in Table 46 and Table 47. For total hip arthroplasties, the cumulative revision rate at 1- year was 1.29%, rising to 2.99% at 5-year and 4.57% at 9-years after the primary surgery. For partial hip arthroplasties, the cumulative revision rates were 1.23%, 2.11%, and 2.38% at 1-, 5- and 9-years after the primary surgery, respectively.

Risk of revision both for patients undergoing total and partial hip arthroplasty was adjusted by sex, age, and comorbidities. As observed for knee arthroplasties, a higher risk of revision was observed in groups of younger patients. The Cox model fitted for total hip arthroplasties revealed a significant statistical risk of revision for patients under 65 years, patients between 65-75 years, and patients between 75-85 years compared to those aged 85 years or over (HR=2.87 (CI 95%: 1.47-5.61); HR=2.38 (CI 95%: 1.22-4.65); HR= 1.96 (CI 95%: 1.00-3.84); data not shown). A similar trend was observed for partial hip arthroplasties for patients aged 65-75 years and 75-85 years (HR=2.81 (CI 95%: 1.78-4.45; HR=1.63 (CI 95%: 1.21-2.20); data not shown).

The cumulative revision rate by fixation technique (without adjustment) for total hip arthroplasties revealed relatively similar revision rates at different follow-up times. Cementless prostheses were those with the highest revision rates at 5- and 9-year of follow-up (3.11% and 5.05%, respectively) (Table 48, Figure 24). For partial hip arthroplasties, the cumulative revision rate for cemented prosthesis was 1.21% at 1-year follow up, and rose to 1.78% and 1.92% at 3- and 5-year follow-up. For cementless prostheses the figures were 1.18%, 1.86% and 2.32% at 1-, 3-, and 5-year of follow-up (Table 49). Risk of revision both for total and partial hip arthroplasties adjusted by sex, age and comorbidities and fixation technique revealed no differences in the risk of revision, age being the only factor associated with revision risk (data not shown).

Table 50 shows the cumulative probabilities of revision per friction device for total hip arthroplasties. Metal on metal prostheses showed revision rates of 1.98%, 3.61% and 5.49% at 1-, 3- and 5- years of follow-up, whereas metal-polyethylene prosthesis presented slightly better outcomes: 1.84%, 2.62% and 3.38%, respectively. The Cox model considering friction device (Metal on Polyethylene, Metal on Metal, Ceramic on Ceramic and Ceramic on Polyethylene) and adjusted by age, sex and comorbidities revealed an increased revision rate for Metal on Metal prosthesis in comparison to Metal on Polyethylene prosthesis (HR=1.43 CI 95%: 1.06-1.93; data not shown).

We estimated the cumulative incidence of revision considering death as a competitive event, both for partial and total hip arthroplasties. Figure 25 shows two functions: one with the cumulative risk of revision adjusted by age, sex, comorbidity and patient death probability, and the other adjusted by the same variables without considering the competitive risk of death. For total hip arthroplasties, both curves are quite similar, the cumulative incidence function without adjustment by competitive risk of death slowly growing apart from the other curve. The cumulative incidence of revision, adjusted by age, sex, and comorbidity at 5 years was 3.20% and considering competitive risk of death, 3.15%. However, for partial knee arthroplasties differences are greater, especially after 2-3 years of follow-up. The cumulative incidence of revision, adjusted by age, sex, and comorbidity at 5 years was 2.36% and considering competitive risk of death 2.05%.

Table 46. Estimated cumulative revision rate after a primary total hip arthroplasty. Source: RACat-MBDSHD-RCA

| Time | Primary | Revisions | Cumulative Revision Rate (%) | CI 95% | |
|---------|---------|-----------|------------------------------|--------|------|
| 30 days | 20,654 | 79 | 0.38 | 0.30 | 0.47 |
| 90 days | 20,106 | 97 | 0.85 | 0.73 | 0.99 |
| 1 year | 18,253 | 85 | 1.29 | 1.14 | 1.45 |
| 3 years | 12,378 | 142 | 2.18 | 1.98 | 2.40 |
| 5 years | 6,600 | 80 | 2.99 | 2.72 | 3.28 |
| 7 years | 2,775 | 42 | 3.88 | 3.51 | 4.28 |
| 9 years | 306 | 12 | 4.57 | 4.03 | 5.17 |

CI 95%: 95% confidence interval. RACat: Catalan Arthroplasty Register

Table 47. Estimated cumulative revision rate after a primary partial hip arthroplasty. Source: RACat- MBD-SHD-RCA

| Time | Primary | Revisions | Cumulative Probability of revision | CI 95% | |
|---------|------------|-----------|------------------------------------|--------|------|
| 30 days | 10,949 | 46 | 0.42 | 0.31 | 0.56 |
| 90 days | 10,684 | 57 | 0.94 | 0.78 | 1.14 |
| 1 year | 9,717 | 30 | 1.23 | 1.04 | 1.46 |
| 3 years | 6,319 | 51 | 1.82 | 1.57 | 2.10 |
| 5 years | 2,969 | 15 | 2.11 | 1.83 | 2.43 |
| 7 years | 1,154 | 1 | 2.18 | 1.87 | 2.54 |
| 9 years | <i>114</i> | <i>1</i> | 2.38 | 1.92 | 2.95 |

CI 95%: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

Table 48. Estimated cumulative revision rate after a primary total hip arthroplasty by fixation technique.
Source: RACat- MBDSHD-RCA

| Time | Primary | Revisions | Cumulative Revision Rate (%) | CI 95% | |
|-----------------------|------------|-----------|------------------------------|--------|------|
| Cemented | | | | | |
| 30 days | 2,049 | 12 | 0.58 | 0.33 | 1.02 |
| 90 days | 2,000 | 12 | 1.17 | 0.78 | 1.74 |
| 1 year | 1,870 | 6 | 1.47 | 1.03 | 2.10 |
| 3 years | 1,258 | 11 | 2.12 | 1.56 | 2.87 |
| 5 years | 730 | 6 | 2.69 | 2.00 | 3.60 |
| 7 years | 380 | 3 | 3.25 | 2.38 | 4.44 |
| 9 years | 66 | 0 | . | . | . |
| Cementless | | | | | |
| 30 days | 13,563 | 45 | 0.33 | 0.25 | 0.44 |
| 90 days | 13,167 | 67 | 0.83 | 0.69 | 0.99 |
| 1 year | 11,826 | 56 | 1.27 | 1.09 | 1.47 |
| 3 years | 7,894 | 89 | 2.12 | 1.88 | 2.40 |
| 5 years | 4,039 | 62 | 3.11 | 2.77 | 3.49 |
| 7 years | 1,553 | 30 | 4.17 | 3.68 | 4.73 |
| 9 years | <i>158</i> | 8 | 5.05 | 4.28 | 5.95 |
| Hybrid | | | | | |
| 30 days | 4,387 | 19 | 0.43 | 0.27 | 0.67 |
| 90 days | 4,292 | 17 | 0.82 | 0.59 | 1.13 |
| 1 year | 3,953 | 20 | 1.30 | 1.00 | 1.68 |
| 3 years | 2,775 | 38 | 2.42 | 1.98 | 2.96 |
| 5 years | 1,583 | 9 | 2.85 | 2.34 | 3.47 |
| 7 years | 770 | 9 | 3.60 | 2.92 | 4.43 |
| 9 years | 71 | 4 | . | . | . |
| Inverse Hybrid | | | | | |
| 30 days | 636 | 3 | 0.47 | 0.15 | 1.45 |
| 90 days | 630 | 1 | 0.63 | 0.24 | 1.66 |
| 1 year | 593 | 3 | 1.10 | 0.53 | 2.30 |
| 3 years | 449 | 4 | 1.86 | 1.03 | 3.34 |
| 5 years | 250 | 3 | 2.69 | 1.56 | 4.60 |
| 7 years | 75 | 0 | . | . | . |
| 9 years | 13 | 0 | . | . | . |

CI 95%: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

Table 49. Estimated cumulative revision rate after a primary partial hip arthroplasty by fixation technique. Source: RACat

| Time | Primary | Revision | Cumulative Revision Rate (%) | IC95% |
|-------------------|---------|----------|------------------------------|-----------|
| Cemented | | | | |
| 30 days | 5,420 | 24 | 0.44 | 0.30 0.66 |
| 90 days | 5,286 | 32 | 1.03 | 0.79 1.34 |
| 1 year | 4,839 | 13 | 1.28 | 1.01 1.62 |
| 3 years | 3,306 | 22 | 1.78 | 1.45 2.18 |
| 5 years | 1,531 | 4 | 1.92 | 1.57 2.36 |
| 7 years | 579 | 1 | 2.06 | 1.64 2.60 |
| 9 years | 67 | 1 | . | . . |
| Cementless | | | | |
| 30 days | 5,527 | 22 | 0.40 | 0.26 0.60 |
| 90 days | 5,396 | 25 | 0.85 | 0.64 1.13 |
| 1 year | 4,877 | 17 | 1.18 | 0.92 1.50 |
| 3 years | 3,013 | 29 | 1.86 | 1.52 2.28 |
| 5 years | 1,438 | 11 | 2.32 | 1.90 2.84 |
| 7 years | 576 | 0 | 2.32 | 1.90 2.84 |
| 9 years | 48 | 0 | . | . . |

CI 95%: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

Figure 24. Cumulative risk of revision after a total hip arthroplasty by fixation technique. Source: RACat-MBDSHD-RCA




Table 50. Estimated cumulative revision rate after a primary total conventional hip arthroplasty per friction device. Source: RACat



| Time | Primary | Revision | Cumulative Revision Rate (%) | CI 95% | |
|--------------------------------|---------|----------|------------------------------|--------|-------|
| Metal on Metal | | | | | |
| 30 days | 970 | 6 | 0.61 | 0.27 | 1.35 |
| 90 days | 936 | 9 | 1.54 | 0.93 | 2.54 |
| 1 year | 864 | 4 | 1.98 | 1.27 | 3.09 |
| 3 years | 687 | 13 | 3.61 | 2.56 | 5.08 |
| 5 years | 420 | 11 | 5.49 | 4.06 | 7.4 |
| 7 years | 134 | 6 | 7.42 | 5.48 | 10.01 |
| 9 years | 12 | 1 | | | |
| Metal on polyethylene | | | | | |
| 30 days | 13448 | 113 | 0.83 | 0.69 | 1 |
| 90 days | 13139 | 83 | 1.45 | 1.26 | 1.66 |
| 1 year | 12111 | 50 | 1.84 | 1.62 | 2.08 |
| 3 years | 8489 | 85 | 2.62 | 2.35 | 2.92 |
| 5 years | 4761 | 53 | 3.38 | 3.05 | 3.74 |
| 7 years | 2109 | 30 | 4.25 | 3.8 | 4.74 |
| 9 years | 219 | 7 | 4.72 | 4.16 | 5.35 |
| Ceramic on Polyethylene | | | | | |
| 30 days | 2413 | 7 | 0.29 | 0.14 | 0.6 |
| 90 days | 2343 | 2 | 0.37 | 0.19 | 0.71 |
| 1 year | 2096 | 13 | 0.95 | 0.63 | 1.45 |
| 3 years | 1401 | 18 | 1.94 | 1.42 | 2.64 |
| 5 years | 760 | 7 | 2.58 | 1.91 | 3.47 |
| 7 years | 318 | 3 | 3.14 | 2.28 | 4.32 |
| 9 years | 52 | 4 | | | |
| Ceramic on Ceramic | | | | | |
| 30 days | 1664 | 5 | 0.3 | 0.12 | 0.71 |
| 90 days | 1617 | 3 | 0.48 | 0.24 | 0.96 |
| 1 year | 1456 | 8 | 0.99 | 0.61 | 1.62 |
| 3 years | 763 | 13 | 2.06 | 1.42 | 2.97 |
| 5 years | 282 | 6 | 3.19 | 2.19 | 4.64 |
| 7 years | 112 | 3 | 4.78 | 3.03 | 7.51 |
| 9 years | 26 | 0 | | | |

CI 95%: 95% confidence interval. RACat: Catalan Arthroplasty Register

Estimates in *italics* indicate that fewer of 250 remain at risk of revision at the time shown. Blank cells indicate that the number at risk at the time shown is fewer than 100 cases. In that case, estimates were not calculated since they are very unreliable.

The number of patients at risk for the other fixation categories was <100; the estimates were not calculated since they were very unreliable.

Figure 25. Cumulative incidence of revision after a partial hip arthroplasty adjusted by competitive risk of death, age and sex. A) Total hip arthroplasty; B) Partial hip arthroplasty. Source RACat-RCA



6. DISCUSSION

The results presented in this report provide an overview of the continued improvement of the Catalan Arthroplasty Register (RACat) regarding the volume and quality of the data. We also report the characteristics of patients undergoing hip and knee arthroplasty, as well as implant characteristics and prosthesis survival considering the procedures carried out between 2005 and 2014 in Catalonia.

6.1. Quality of the data: volume of interventions and completeness

Following a steady rise in the number of hip and knee cases reported to the RACat as of 2005, the data demonstrates a decrease between 2010 and 2011. This decrease was linked to a reorganisation of hospital care carried out in Catalonia that year, given that once this had finished in 2012, the number of interventions being reported to the register started to gradually rise again.

However, the increase in the number of arthroplasties in the last year was not reflected in the data sent to the RACat. This was due to the fact that, at the end of 2013, hospitals had to temporarily stop their data notifications in order to adapt their IT systems to the technical requirements of the Catalan Health Service (CHS), which included a new web services communication process. This is the reason why RACat completeness values also decreased during the latter period (2013-2014). Nevertheless, they can be expected to improve as a result of the satisfactory resolution of this circumstance.

Even with the improvement of completeness and the quality of the data, further strategies must be designed to allow even greater improvement to continue. The participation of private centres will be essential in order to have a complete picture of the healthcare assistance activity in Catalonia.

6.2. Description of patients and the healthcare process

Confirming trends observed in previous reports and in other countries^{3,4,10-12}, the current results show that both knee and hip arthroplasties were more frequently performed among women than among men, osteoarthritis being by far the most prevalent reason for intervention in total hip and both total and partial knee arthroplasties. Meanwhile, mechanical complications of orthopaedic devices were the main reason for revision, a situation that had already been observed in other European registers, which reported aseptic loosening as the main cause for revision³. In the RACat, aseptic loosening was analysed together with other diagnoses such as dislocation and contact surface wear, under the concept of mechanical complications.

Results pointed out that the majority of patients undergoing hip and knee arthroplasties presented one or more comorbidities. This may be related to increased life expectancy and to improvements in surgical techniques, which have made it possible to operate on older patients, who tend to have a higher number of comorbidities. However, some specific comorbidities such as obesity, renal failure and neurological disorders, have been highlighted as obstacles in the

recovery process, since these patients will be less likely to increase their physical activity after elective surgery¹³.

Regarding the type of discharge from hospital, the most common destination was home. Although this appears to be better in terms of early recovery¹⁴, patient satisfaction and health-care costs, there is no clear evidence regarding the impact it might have on the long-term recovery process¹⁵. For this reason, it may be interesting for future studies to look at potential readmissions of patients per their discharge type, in order to obtain more evidence on the healthcare quality of this practice after an arthroplasty.

6.3. Characteristics of primary arthroplasties and implant models

Our findings revealed significant variability in the number of prosthesis models used depending on the hospital, both for knee and hip arthroplasties. Some studies suggested that using a high number of models could be a risk factor for revision¹⁶. Based on the results of the Swedish register, this group recommends limiting the number of models in use, with expectations of reducing the risk of revision. This measure, if confirmed, could be of interest for Catalonia, given the wide variability observed in our hospitals. Further analyses aimed at exploring the specific results of implant models would facilitate the decision-making process for professionals based on reliable clinical evidence. Some initiatives, such as the Orthopaedic Data Evaluation Panel (ODEP)¹⁷ in the UK, or the Dutch Hip Task Force (DHTF)¹⁸ in The Netherlands, provide on-going assessment of hip and knee implants presenting a benchmark rating for implant survival.

In accordance with trends observed in the latter period, the most common knee procedure was total knee arthroplasty. Although the number of partial arthroplasties has increased slightly during the most recent periods, their impact on overall knee procedures remains low. In the case of primary hip arthroplasties, the number of partial procedures has increased. For this kind of surgery, we observed a decreasing trend over time on the use of unipolar monoblock and an increase of bipolar prosthesis. This pattern was also observed in Australia¹⁹ but contrasts with trends reported in Sweden¹¹.

Regarding the type of fixation, RACat recorded an increase in the volume of primary cemented knee arthroplasties, confirming trends observed in latter periods. On the contrary, most of total hip arthroplasties were cementless, this type of fixation showing an increasing trend over time, especially among young patients. This pattern is consistent with trends reported in other countries like Norway²⁰, Sweden¹¹, New Zealand²¹, and the United Kingdom²². Likewise, the use of metal on metal as bearing surface has decreased over time, whereas metal-polyethylene is increasingly used.

6.4. Implant survival

In general, both knee and hip arthroplasties presented good implant survival, both in the short, medium, and long-term. Total knee arthroplasties showed better survival outcomes than partial replacements. In particular, our findings show a 9-year revision rate less than 4% for cemented and hybrid prosthesis, this data being in accordance with that reported in other countries²². Unicompartamental replacements showed worse survival results compared to total knee replacements with the chance of revision at each estimation time point being around twice that of a

total knee replacement. It has to be considered that this surgery has different indications than total knee arthroplasty, being more commonly performed among younger patients. This may be related to milder disease in these patients, or the desire to delay a total knee replacement for as long as possible. Moreover, younger patients may also be more active and are more likely to present better health conditions, which puts more strain on their implants and allows reoperation if a complication or failed prosthesis occurs.

Following trends observed in previous reports, partial hip arthroplasties presented better survival outcomes than total hip procedures. Again, differences in patient's age may explain the different results, given that partial arthroplasties are more commonly performed in older patients, as an indication of femoral neck fracture.

As expected, metal on metal implants showed significantly higher revision rates in comparison to other friction devices, confirming in our setting the unsatisfactory results of these bearing devices in the medium- and long-term reported in the literature²³. This situation calls for close surveillance of patients with this kind of prosthesis and for monitoring the degree of accomplishment of recommendations issued by public bodies. In parallel, current data points out that ceramic–metal implants presented the lowest failure rates, followed by metal-polyethylene implants.

6.5. Strengths and weaknesses

After 10 years of implementation of the RACat, current data provides us with an overall picture of the performance of hip and knee arthroplasties in Catalonia as well as robust estimators on implant utilization and survival outcomes. However, we would like to point out certain limitations that should be taken into consideration when interpreting the results presented herein. Firstly, the volume of cases in some subgroups remains low, especially in the survival analyses. This situation hampers the calculation of risk of revision in the medium- and long-term. As a result, long-term survival outcomes for some figures should be interpreted with caution.

Secondly, we wish to point out the high number of cases in which the type of fixation was not specified, most notably in the earlier years of the register. During this period, the emphasis was on the implementation of the register, without as much emphasis being placed on the quality of the data as there is now. It is hoped that in coming years the percentage of cases not reported to the RACat will gradually decrease, as has been the trend to date. Finally, as we have already mentioned above, the information on cause for revision was obtained by merging our database with the MBDSHD. As pointed out, there are some deficiencies in the MBDSHD coding that hamper the level of detail of this variable. We do believe that with the implementation of the surgical form, we will overcome this limitation, given that this form will allow us to collect more accurate information regarding this variable along with other clinical information.

Despite the limitations mentioned above and the difficulty and high cost involved in implementing and maintaining a register of this kind, we can assert that the RACat has consolidated itself as a powerful tool for analysing the medium- and long-term effectiveness of hip and knee arthroplasties, for studying the variability of clinical practice, and for post-market monitoring.

7. FINAL REMARKS AND NEXT STEPS

7.1. Latest achievements

Year after year we stress our commitment to improve the quality of the Register and to publicize the results and the potential of this tool for managers, surgeons and healthcare professionals. In this line, we would like to highlight some of the activities and products that have been released over the last year.

Commemoration of the 10th anniversary of the RACat. A multidisciplinary, informative session was organised in conjunction with the Catalan Society of Orthopaedic Surgery and Traumatology (SCCOT). Experts from different fields of health care, management, research, industry, and administration were invited, with the aim of pooling knowledge and perspectives on the scope and implications of the RACat. These different points of view, along with the patient's vision, were recorded in an educational video, which is available on our website: http://aquas.gencat.cat/ca/projectes/mes_projectes/qualitat_atencio_sanitaria/racat/

Atlas of Variation of Knee and Hip Arthroplasties in Catalonia. As part of a wider project lead by the Agency for Health Quality and Assessment (AQuAS), Atlases of Variations in Knee and Hip Arthroplasties in Catalonia were presented using an interactive format. This new format, which uses the InstantAtlas Dynamic Report Platform, allows us to view data from different basic healthcare areas in Catalonia and to benchmark the performance of both surgeries over different time periods. These are also available on our website: http://aquas.gencat.cat/ca/projectes/atles-de-variacions-i-de-qualitat-del-siscat/atles_variacions/atles_finalitzats/atles_genoll/atles_digital_artroplasties_genoll_maluc/

Research. The RACat is also involved in several research projects lead by clinicians, epidemiologists, and healthcare professionals, which explore different issues related to arthroplasties performance and its impact on the healthcare system.

7.2. Next steps

For the upcoming years, we face new challenges that we are sure will contribute to improving the quality of the register and will promote the visualization and utilization of the RACat.

Implementation of the surgical form. The inclusion of variables from the surgical form will substantially increase the information recorded in the register (mainly clinical information), enriching the current dataset and enhancing the possibility to refine some analyses. Although some hospitals are currently reporting the variables of this form, there is still room for improvement in its complete implementation in Catalonia.

Data from private centres. The current scope of the register includes hospitals run by the Integrated Healthcare System of Public Healthcare Network of Catalonia (SISCAT). Widening the

scope to private and public centres will enable us to get a complete picture of the arthroplasties carried out in Catalonia (reaching 100% coverage) and will enable the monitoring and surveillance of all prosthesis used in our country.

Continuous improvement of data quality. We are still making efforts to improve the quality of the information sent by centres. We periodically perform data quality reviews, schedule face-to-face meetings with centres, and promote close collaboration with professionals. However, the best way to promote and encourage participation is to make the register more useful for professionals according to their needs. This is one of the main challenges we face.

Linking the RACat with different databases. In recent years, we have carried out *ad-hoc* linkages with databases in order to answer specific research questions arising out of different projects. This has enabled collaborations with researchers from other disciplines and has enhanced the possibilities of data exploitation. We hope to consolidate past collaborations and to initiate new projects.

Promote dissemination. Participation in scientific forums and international networks will promote the visualization and dissemination of the RACat both inside Catalonia and beyond our borders. In this line, we have planned to improve the information displayed on our website to make it more attractive, visual, and current.

8. REFERENCES

1. Guerrero-Ludueña R, Comas M, Espallargues M, Coll M, Pons M, Sabatés S, et al. Predicting the Burden of Revision Knee Arthroplasty: Simulation of a 20-Year Horizon. *Value Health*. 2016;19(5):680-7.
2. Serra-Sutton V, Allepuz A, Espallargues M, Labek G, Pons JM. Arthroplasty registers: a review of international experiences. *Int J Technol Assess Health Care*. 2009;25(1):63-72.
3. Delaunay C. Registries in orthopaedics. *Orthop Traumatol Surg Res*. 2015;101(1 Suppl):S69-75.
4. Serra-Sutton V, Tebé C, Martínez O, García-Altés A, Espallargues M; in name of the Advisory Committee and the Management of the RACat. *Registre d'Artroplàsties de Catalunya. Segon informe dades 2005-2010*. Barcelona: Agència d'Informació, Avaluació i Qualitat en Salut. Servei Català de la Salut. Departament de Salut. Generalitat de Catalunya; 2012.
5. Martí-Valls J, Alonso J, Lamarca R, Pinto J, Auleda J, Girvent R, et al. Efectividad y costes de la Intervención de prótesis total de cadera en siete hospitales de Cataluña. *Med Clin (Barc)*. 2000;114 Suppl 2:34-9.
6. *Registre d'Artroplàsties de Catalunya (RACat). Documentació tècnica [Internet]*. Barcelona: Agència de Qualitat i Avaluació Sanitàries de Catalunya; 2012 [cited January 2017]. Available from: http://aquas.gencat.cat/ca/projectes/mes_projectes/qualitat_atencio_sanitaria/racat/documentacio_tecnica/.
7. *Australian Orthopaedic Association National Joint Replacement Registry [Internet]*. [cited January 2017]. Available from: <https://aoanjrr.sahmri.com/>.
8. Cox DR, Oakes D. *Analysis of Survival Data*. New York: Chapman & Hall; 1984.
9. Fine J, Gray R. A proportional hazards model for the subdistribution of a competing risk. *J Am Statist Assoc*. 1998;94(446):496-509.
10. *Insight into Quality & Safety. Annual Report of the Dutch Arthroplasty Register. LROI Report*. 's-Hertogenbosch (The Netherlands): Dutch Arthroplasty Register (LROI); 2013.
11. *The Swedish Hip Arthroplasty Register. Annual Report 2014*.
12. *Swedish Knee arthroplasty register. Annual report 2015*.
13. Paxton E, Torres A, Love RM, Barber TC, Sheth DS, Inacio MSC. Total joint replacement: A multiple risk factor analysis of physical activity level 1-2 years postoperatively. *Acta Orthop*. 2016 Jul;87 Suppl 1:44-9.
14. Keswani A, Tasi MC, Fields A, Lovy AJ, Moucha CS, Bozic KJ. Discharge Destination After Total Joint Arthroplasty: An Analysis of Postdischarge Outcomes, Placement Risk Factors, and Recent Trends. *J Arthroplasty*. 2016;31(6):1155-62.
15. Shepperd S, Doll H, Broad J, Gladman J, Iliffe S, Langhorne P, et al. Early discharge hospital at home. *Cochrane Database Syst Rev*. 2009;1:CD000356.

16. Heberts P, Malchau H. Long-term registration has improved the quality of hip replacement: a review of the Swedish THR Register comparing 160,000 cases. *Acta Orthop Scand.* 2000;71:111-21.
17. Orthopaedic Data Evaluation Panel (ODEP). Available from: <http://www.odep.org.uk/>.
18. Poolman RW, Verhaar JA, Schreurs BW, Bom LP, Nelissen RG, Koot HW, et al. Finding the right hip implant for patient and surgeon: the Dutch strategy--empowering patients. *Hip Int.* 2015;25(2):131-7.
19. Australian Orthopaedic Association National Joint replacement Registry. Annual Report 2015.
20. Norwegian National Advisory Unit on Arthroplasty and Hip Fractures. Annual Report 2015.
21. The New Zealand Joint Registry. Sixteen year report, October 2015.
22. National Joint Registry for England, Wales, Northern Ireland and the Isle of Man. 12th Annual report 2015.
23. Pijls BG, Meessen JMTA, Schoones JW, Fiocco M, van der Heide HJL, Sedrakyan A, Nelissen RGHH. Increased Mortality in Metal-on-Metal versus Non-Metal-on-Metal Primary Total Hip Arthroplasty at 10 Years and Longer Follow-Up: A Systematic Review and Meta-Analysis. *PLoS One.* 2016;11(6):e0156051.



Founding Member



Corporate Member

ciberesp



